

**SPECIFICATIONS OF FLUORESCENT ILLUMINATOR TO VIEW
ON A STAND 4 FILM OF 14" X 17" LED VIEW BOX**

Technical Requirements:-

- It should have viewing area to accommodate four 14'x17' films (Four 14'x17 films viewing slots)
- The light source should be LED lamps, having life span more than 100000 Hours.
- It should provide flicker free & homogeneous illumination.
- It should have on/off button.
- It should have brightness control with adjustment range with 80% or more in every slot.
- Thickness should be 2 inch or less.
- There should be easy & firm insertion & removal of film.
- The device should have external fuses for protection.
- The power supply input should be of AC 220-240V, with direct connectivity without any external adapter.

Other Requirements:-

- The manufacturer company should be ISO certified.
- Three years guarantee including all accessories from the date of installation.
- The company should have a local Service centre.

The company should have proven track record in Govt. sector list of previous installation to be submitted.

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur