

SPECIFICATIONS OF SINGLE CHANNEL ECG MACHINE


(Fully automatic, microprocessor controlled, Single-channel resting electrocardiograph, mains and battery operated)

Standard unit including :

1. 12 lead simultaneous ECG data acquisition
2. High-resolution built-in thermal printer
3. Sensitivity : 2.5, 5, 10, 20 mm/mV
4. Should have Membrane keypad for operating
5. Printout on 50 mm or more Z fold paper or roll type paper
6. Battery should last minimum up to 200 recordings
7. ECG measurements, Copy Function
8. Auto & Manual Mode
9. Printing Mode : Auto simultaneous acquisition, with parameter calculation and Manual Real Time
Should be supplied with USB & Software for storage of ECG on Computer.

Standard accessories included :

1. Patient Cable: 10 lead
2. Recording paper - 1
3. Clip-on Electrode - 4
4. Jelly Bottle - 1
5. Chest Electrode with Bulb - 6
6. Earth Cable - 1
7. Warranty Card
8. Operating manual



**Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur**