


SPECIFICATIONS OF SEROLOGICAL WATER BATH
(ISI Mark 6593-1978)

Water bath with double walled (outer chamber made of mild steel sheet with finished powder coated paint & inner chamber made of stainless steel); gap between the walls should be filled with glass wool. The water bath should be provided with a drain plug to facilitate easy emptying & cleaning of inner chamber. Pyramid call shape Cover. Provided with immersion heating elements Digital Display temperature indicator cum controller.

Size (Inner chamber) : 330 mm X 300mm X 150mm for 4 racks 1.5 kw



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur 