## SPECIFICATIONS OF TRANSPORT INCUBATOR, BASIC, WITH BATTERY AND O<sub>2</sub>, W/O VENTILATOR

- Double wall transparent canopy with mattress, mount on stretcher
- Front and head access door, slide-out mattress tray
- With baby restraining straps
- Warm air circulation system
- Bacterial filter to remove air born particles
- Incubator air temperature monitoring and servo control: 25 to 38 C, increments 0.1C
- Digital displays outside shows air temperature
- Two 5 L integrated aluminum oxygen cylinders, regulator and flow meter
- Audiovisual alarms: high/low air temperature, temperature sensor failure, power failure an low battery
- Construction dismantleable allows frequent washing and disinfection of the incubator
- Battery and AC supported
- Power requirements: 220 V/50 Hz and internal re-chargeable batteries (autonomy approx 3 hrs, automatic recharge)
- Power consumption: 200 W
- Device is produced by ISO 9001 certified manufacturer
- Device is safety certified according ISO & CE (93/42)/ FDA 510k.

## **Supplied with:**

- 1 x spare air temperature probe
- 1 x spare rechargeable battery
- 2 x empty 5 L aluminum oxygen cylinders
- 1 x spare set of fuses
- User manual with trouble shooting guidance, in English
- Technical manual with maintenance and first line technical intervention instructions, in English
- List of priced accessories
- List of priced spare parts
- List with name and address of technical service providers in India
- Training and installation at end-user site
- Proposal for full service 1 year & battery two years Guarantee and 3 years Comprehensive Annual Maintenance Charges (CAMC) covering (i) 2 preventive maintenances per year,
- (ii) on-call technical interventions, spare parts and travel

Executive Director (EPM)

Medical Services Corporation

Rajasthan Medical Services Corporation Ltd.,

Jaipur

L