

**List of approved and not approved reagents kit for tender in RMSC (2013)**

S.No.	Item	Approved	Not approved	Reason
1	S. Creatinine	1. Transasia 2. Randox		
2	Total Cholestrol	1. Transasia	1. Randox	Linearity not upto the mark
3	Amylase	1. span 2. Agappe 3. Transasia	1. Randox	Linearity not upto the mark
4	Uric Acid	1. Transasia 2. Randox 3. Agappe		
5	<del>Electrolyte</del>	<del>1. Randox</del> <del>2. Transasia</del>		
6	S.LDH		1. Randox 2. Transasia	Linearity not upto the mark Method not upto the Mark

(Dr. Harjeet Singh)  
Asst. Prof, Biochemistry

(Dr. Rati Mathur)  
Professor, Bio Chemistry

(Dr. S.K. Vardey)  
PHOD, Biochemistry