



Rajasthan Medical Services Corporation Limited

Gandhi Block, SwasthyaBhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 ,2228064 Fax No. 0141-2228065 Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E_mail :edprocurement@gmail.com, andrmsc@nic.in

Ref. No.:- F.02(289)/RMSCL/Proc/PPE kits/NIB-12/2020/963


Dated: 01.06.2020

Corrigendum – III

Subject:- Amended condition / technical specification/ Bid Security


Ref.:- NIB No. F.02(289)/RMSCL/Proc/PPE kits/NIB-12/2020/933 Dated:23.05.2020 (Technical bid opening due on dated –05.06.2020)

S. No	Amended condition / technical specification / Bid Security (clause no.).																														
1.	<p>1. Regarding bid security (clause 5(a), 9):- It was decided to accept bid security in the form of bank guarantee also.</p> <p>2. Item wise Bid security:- In the annexure VI technical specification, as each item is a separate entity for the purpose of technical and financial evaluation, therefore it was decided to indicate item wise amount of bid security</p> <table border="1"> <thead> <tr> <th>S. No.</th> <th>Item code</th> <th>Tender quantity</th> <th>Bid security</th> <th>Bid security (For MSME)</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>PPE Kit 1A (Non sterile)</td> <td>100000</td> <td>7 lacs</td> <td>1.75 lacs</td> </tr> <tr> <td>2</td> <td>PPE Kit 1B (sterile)</td> <td>100000</td> <td>7 lacs</td> <td>1.75 lacs</td> </tr> <tr> <td>3</td> <td>PPE Kit II</td> <td>800000</td> <td>54.60 lacs</td> <td>13.65 lacs</td> </tr> <tr> <td>4</td> <td>PPE Kit III</td> <td>200000</td> <td>60,000</td> <td>15,000</td> </tr> <tr> <td>5</td> <td>PPE Kit IV</td> <td>200000</td> <td>80,000</td> <td>20,000</td> </tr> </tbody> </table> <p>3. Regarding technical specifications provided in Annexure VI of bid document. following are clarified-</p> <p>For item code PPE Kit 1A, PPE Kit 1B, PPE Kit III</p> <ol style="list-style-type: none"> All documents regarding SITRA/ DRDO certification for fabric as well as for entire PPE should be submitted along with technical bid. The successful bidder while supplying every lot will also be required to submit a SITRA/DRDO certificate establishing that he has used the material certified by SITRA/DRDO along with invoice. He would be also being required to submit this certification for the entire PPE kit. For PPE kit of sterile variety, a test certificate for sterility should also be submitted. The test report should be from NABL accredited Lab. For item code PPE Kit III shoe cover the quantity will be 200000 Pairs. Bidders need to submit samples of the quoted product/ products within three days of the technical bid submission. <p>4. Drawing and design specifications of Face cover:-</p>	S. No.	Item code	Tender quantity	Bid security	Bid security (For MSME)	1	PPE Kit 1A (Non sterile)	100000	7 lacs	1.75 lacs	2	PPE Kit 1B (sterile)	100000	7 lacs	1.75 lacs	3	PPE Kit II	800000	54.60 lacs	13.65 lacs	4	PPE Kit III	200000	60,000	15,000	5	PPE Kit IV	200000	80,000	20,000
S. No.	Item code	Tender quantity	Bid security	Bid security (For MSME)																											
1	PPE Kit 1A (Non sterile)	100000	7 lacs	1.75 lacs																											
2	PPE Kit 1B (sterile)	100000	7 lacs	1.75 lacs																											
3	PPE Kit II	800000	54.60 lacs	13.65 lacs																											
4	PPE Kit III	200000	60,000	15,000																											
5	PPE Kit IV	200000	80,000	20,000																											

S. No	Amended condition / technical specification / Bid Security (clause no.).
	 <p data-bbox="274 651 1255 689">In the face shield the transparent window will have following specifications –</p> <p data-bbox="217 703 788 741">Overhead Projector / Clear transparency film</p> <ol data-bbox="289 741 751 779" style="list-style-type: none"> <li data-bbox="289 741 751 779">1. Wrinkle free, with clear visibility <p data-bbox="289 815 514 846">Size of sheet/film</p> <p data-bbox="278 853 468 884">Length: 14 cm</p> <p data-bbox="278 891 448 922">Width: 20cm</p> <p data-bbox="278 929 575 960">Micron: 200 Minimum</p> <p data-bbox="263 994 1379 1070">Fabric of face cover for PPE Kit 1A, PPE Kit 1B, and PPE Kit II should be as specified in specifications in Annexure VI of bid document.</p>

Note:-

- It may be noted that if any further amendments are issued then a corrigendum will be published and informed.
- Rest of the terms and conditions will remain the same.


 Executive Director (Proc.)
 RMSCL