



Rajasthan Medical Services Corporation Limited

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 , 2228064 Fax No. 0141-2228065 Website: <http://rmsc.healtht.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E_mail : edprmsc@nic.in,

Ref. No.: F.02(316)/RMSCL/Proc/PPE kits/NIB-07/2021/628

Dated:- 26.4.2021

Corrigendum – II

Subject:- Amendments in Date, BoQ technical specification/conditions

Ref. No.: NIB No. F.02(316)/RMSCL/Proc/PPE kits/NIB-07/2021/506 Dated:-30.03.2021


(Technical Bid Open on 27.04.2021)

S. No	Existing technical specification (clause no.)	Amended technical specification (clause no.).												
1	<p>Technical Bid opening date F.02(316)/RMSCL/Proc/PPE kits/NIB-07/2021/506 Dated:-30.03.2021 is 4.00 PM of 27.04.2021.</p> <table border="1"> <tr> <td>Last date and time of submission of online bids</td> <td>27.04.2021 at 3.00 PM</td> </tr> <tr> <td>EMD, Tender fees, RISL fees through challan and Physically</td> <td>27.04.2021 at 3.00 PM</td> </tr> <tr> <td>Date and time of opening of Online technical bids</td> <td>27.04.2021 at 4.00 PM</td> </tr> </table>	Last date and time of submission of online bids	27.04.2021 at 3.00 PM	EMD, Tender fees, RISL fees through challan and Physically	27.04.2021 at 3.00 PM	Date and time of opening of Online technical bids	27.04.2021 at 4.00 PM	<p>Technical Bid opening date F.02(316)/RMSCL/Proc/PPE kits/NIB-07/2021/506 Dated:-30.03.2021 is 4.00 PM of 30.04.2021.</p> <table border="1"> <tr> <td>Last date and time of submission of online bids</td> <td>30.04.2021 at 3.00 PM</td> </tr> <tr> <td>EMD, Tender fees, RISL fees through challan and Physically</td> <td>30.04.2021 at 3.00 PM</td> </tr> <tr> <td>Date and time of opening of Online technical bids</td> <td>30.04.2021 at 4.00 PM</td> </tr> </table>	Last date and time of submission of online bids	30.04.2021 at 3.00 PM	EMD, Tender fees, RISL fees through challan and Physically	30.04.2021 at 3.00 PM	Date and time of opening of Online technical bids	30.04.2021 at 4.00 PM
Last date and time of submission of online bids	27.04.2021 at 3.00 PM													
EMD, Tender fees, RISL fees through challan and Physically	27.04.2021 at 3.00 PM													
Date and time of opening of Online technical bids	27.04.2021 at 4.00 PM													
Last date and time of submission of online bids	30.04.2021 at 3.00 PM													
EMD, Tender fees, RISL fees through challan and Physically	30.04.2021 at 3.00 PM													
Date and time of opening of Online technical bids	30.04.2021 at 4.00 PM													
2	<p>Annexure – VI (Specifications of PPE Kits) and BoQ</p> <table border="1"> <thead> <tr> <th>Specifications of PPE Kits</th> <th>Packaging Unit</th> <th>Tender Quantity</th> </tr> </thead> <tbody> <tr> <td> <p>Coverall & Shoe cover should be 95 GSM splash proof. Entire kit should be non-sterile.</p> <ul style="list-style-type: none"> • Face cover in which transparent window in face cover should be wrinkle free with clear visibility and should have tape/ string to tie across chest. • Taping on seams & fold on zipper. • Nitrile powder free gloves. • Bag for disposal of kit. <p>The bidder should submit a test report from NABL Accredited lab in proof of GSM of coverall and shoe cover.</p> </td> <td>Piece comprising all components</td> <td>300000</td> </tr> </tbody> </table>	Specifications of PPE Kits	Packaging Unit	Tender Quantity	<p>Coverall & Shoe cover should be 95 GSM splash proof. Entire kit should be non-sterile.</p> <ul style="list-style-type: none"> • Face cover in which transparent window in face cover should be wrinkle free with clear visibility and should have tape/ string to tie across chest. • Taping on seams & fold on zipper. • Nitrile powder free gloves. • Bag for disposal of kit. <p>The bidder should submit a test report from NABL Accredited lab in proof of GSM of coverall and shoe cover.</p>	Piece comprising all components	300000	<p>Annexure – VI (Specifications of PPE Kits) and BoQ</p> <table border="1"> <thead> <tr> <th>Specifications of PPE Kits</th> <th>Packaging Unit</th> <th>Tender Quantity</th> </tr> </thead> <tbody> <tr> <td> <ul style="list-style-type: none"> • Coverall fabric-Nonwoven with breathable fabric & lamination. • Material should be splash-proof. • Coverall & shoe cover should be of 90±5 GSM. • Taping on seams & fold on zipper. • Face cover & with transparent window which should be wrinkle free & should have clear visibility and should have tape/string to tie across chest. • Shall not disintegrate, split or tear during use. • Shoe cover with length of leggings upto knee with 2 layers of elastic; one in the ankle area & another in the leg area. • Coveralls to be of be of variable sizes. Medium large & extra large. • Packaging to maintain integrity of the product. • Nitrile powder free gloves to be provided along with the kit. • Should come with appropriate waste bag for disposal of kit. • The bidder should submit a test report from NABL Accredited lab in proof of GSM of coverall and shoe cover. </td> <td>Piece comprising all components</td> <td>500000</td> </tr> </tbody> </table>	Specifications of PPE Kits	Packaging Unit	Tender Quantity	<ul style="list-style-type: none"> • Coverall fabric-Nonwoven with breathable fabric & lamination. • Material should be splash-proof. • Coverall & shoe cover should be of 90±5 GSM. • Taping on seams & fold on zipper. • Face cover & with transparent window which should be wrinkle free & should have clear visibility and should have tape/string to tie across chest. • Shall not disintegrate, split or tear during use. • Shoe cover with length of leggings upto knee with 2 layers of elastic; one in the ankle area & another in the leg area. • Coveralls to be of be of variable sizes. Medium large & extra large. • Packaging to maintain integrity of the product. • Nitrile powder free gloves to be provided along with the kit. • Should come with appropriate waste bag for disposal of kit. • The bidder should submit a test report from NABL Accredited lab in proof of GSM of coverall and shoe cover. 	Piece comprising all components	500000
Specifications of PPE Kits	Packaging Unit	Tender Quantity												
<p>Coverall & Shoe cover should be 95 GSM splash proof. Entire kit should be non-sterile.</p> <ul style="list-style-type: none"> • Face cover in which transparent window in face cover should be wrinkle free with clear visibility and should have tape/ string to tie across chest. • Taping on seams & fold on zipper. • Nitrile powder free gloves. • Bag for disposal of kit. <p>The bidder should submit a test report from NABL Accredited lab in proof of GSM of coverall and shoe cover.</p>	Piece comprising all components	300000												
Specifications of PPE Kits	Packaging Unit	Tender Quantity												
<ul style="list-style-type: none"> • Coverall fabric-Nonwoven with breathable fabric & lamination. • Material should be splash-proof. • Coverall & shoe cover should be of 90±5 GSM. • Taping on seams & fold on zipper. • Face cover & with transparent window which should be wrinkle free & should have clear visibility and should have tape/string to tie across chest. • Shall not disintegrate, split or tear during use. • Shoe cover with length of leggings upto knee with 2 layers of elastic; one in the ankle area & another in the leg area. • Coveralls to be of be of variable sizes. Medium large & extra large. • Packaging to maintain integrity of the product. • Nitrile powder free gloves to be provided along with the kit. • Should come with appropriate waste bag for disposal of kit. • The bidder should submit a test report from NABL Accredited lab in proof of GSM of coverall and shoe cover. 	Piece comprising all components	500000												



Note:-

- It may be noted that if any type of amendments required then further corrigendum will be published and informed.
- Rest terms and conditions will remain the same.


(Shrinivas Meena)
Executive Director (Proc.)
RMSCL

A