



**Rajasthan Medical Services Corporation Limited, Jaipur
D-Block, Swasthya Bhawan, C-Scheme, Jaipur - 302005**

Ph. No. 0141-2223887, Fax No. 0141-2228065
CIN : U24232RJ2011SGC035067

E-Mail – mdrmsc@nic.in; edepmrmsc-rj@nic.in
Website: http://rmsc.health.rajasthan.gov.in

F-8(199) RMSC/EPM/M-2/NIB-473/2019-20/4186 Dated: 04/06/2020

Clarification/Corrigendum/Addendum

Subject:-Revised Technical Specification for NIB No. F-8(199) RMSC/EPM/M-2/NIB-473/2019-20/3550 dated 17.02.2020 for item Falcon Tube (PP Tubes for Centrifuge).

In reference to subject cited above of NIB-473, various representations received from the firms and issues raised by the bidders are being examined authorities and technical committee. The following corrigendum/addendum is issued for inclusion in bid document as below:-

1. Revised Technical Specification of centrifuge tube 50 ml.

1. Sterile Centrifuge tubes 50 ml made of polypropylene, certificate of type of sterilization should be attached with technical bid.
2. Any color labeling area and printed graduation at every 5 ml or less.
3. Walls thick enough to allow centrifugation up to 1500g.
4. Solvent resistant, chemicals compatibility certificate to be documented.
5. DNase/RNase-Free.
6. Should withstand at 15 psi pressure & at 121^o C temperature.
7. Manufacturer should be ISO 9001 Certified.
8. Certified Non-Pyrogenic (Certificate to be attached).
9. Individually packed with manufacturing and expiry date on label.
10. Flat PE screw caps should close with quick 3-4 turns.
11. 10 Nos. samples of item should be submitted along with technical bid.

Please note that all clarifications/amendments/corrigendum in technical specifications/bid conditions/date extension of bid are integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remain the same.

Executive Director (EPM)
RMSCL, Jaipur