



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail: edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmhc.health.rajasthan.gov.in

F-8) RMSC/EPM/M-5/NIB-705/ 560

Date:- 21-10-2022

Clarification/Corrigendum/Addendum

**Subject:-Revised Technical Specifications of the rate contract for 1. Table Top Tube Sealer
2. Dielectric Tube Sealer, Handheld (Portable) UNIT for under NIB No. F-8)
RMSC/EPM/M-5/NIB-705/2022-23/464 Date 21.09.2022**

In Reference to subject cited above and NIB-705, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

Revised Technical Specifications:-

Item- Table Top Tube Sealer

TECHNICAL SPECIFICATIONS	AMENDED TECHNICAL SPECIFICATIONS
15. Bidder should submit USFDA certificate/European CE Certificate from notified body for quoted product (Valid documentation should be submitted in technical bid).	15. Bidder should submit USFDA certificate/European CE Certificate as per applicable device category for quoted product (Valid documentation should be submitted in technical bid).

Item-Dielectric Tube Sealer, Handheld (Portable)

TECHNICAL SPECIFICATIONS	AMENDED TECHNICAL SPECIFICATIONS
2. (b) Bidder should submit USFDA certificate/European CE Certificate from notified body for quoted product (Valid documentation should be submitted in technical bid).	2. (b) Bidder should submit USFDA certificate/European CE Certificate as per applicable device category for quoted product (Valid documentation should be submitted in technical bid).
3. (J) No. of seals per charge should be minimum 300 to 1000 continuous seals from a fully charged battery.	3. (J) No. of seals per charge should be minimum 700 to 1000 continuous seals from a fully charged battery.



Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur