

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrm-sc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rm-sc.nic.in

No. F-8()RMSC/EPM/M-2/2015-16/NIB-164/ 2213

Dated: 9/5/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical Specifications of items and revised bid submission date under NIB No. F-8()RMSC/EPM/M-2/2015-16/NIB-164/1500 dated 28.03.16

In Reference to subject cited above and NIB-164, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

1. Amended Technical Specifications of item Operation Table Hydraulic Minor:-

Technical Characteristics	
1.	Should have stainless steel top grade 304.
2.	Should have heavy duty castor wheel for easy mobility.
3.	Head and Leg Section should be detachable and interchangeable.
4.	Table should have five sections (with kidney bridge)
5.	Durable and leak proof hydraulic pump with heavy pillar fitted in centre of the table.
6.	Base and column covered with stainless steel grade 304.
7.	Should have gear mechanism.
8.	Should have base locking mechanism for rigid locking.
9.	Stainless Steel side rails accept all standard accessories.
Physical Characteristics	
1.	Table Top Dimension : Length: 1900 mm to 2000mm Width : 500mm to 550 mm
2.	Table Elevation : Maximum height from floor : 800mm and above Minimum height from floor : 525mm to 600mm or less
3.	Trendelenburg and reverse Trendelenburg position: 25° - 30° on each way.
4.	Lateral tilt (Left & Right): 15° - 20° on each side.
5.	Sitting position upto 75° or more.
6.	Leg section : Up : 80° - 90° Down: 105° ± 10°
7.	Head section : Up : 45° - 60° Down: 45° - 90°
8.	Should be able to support the weight of the patient upto 200kg
Accessories	
1.	Fluid proof mattress for the complete table in sections - 1 Set.
2.	Shoulder Support with clamps - 1 Pair.
3.	Padded arm rest with clamp - 1 Pair
4.	IV Pole - 1 No.
5.	Body restraining belt - 1 No.
6.	Leg supports with clamp - 1 Pair.
7.	Lateral Supports with clamp - 1 Pair.
8.	Aesthetic screen frame with clamp - 1 No.
Standards and Safety	
1.	Should be CE / BIS approved product.
2.	Manufacturer should have ISO 13485 certificate.
Training and Installation	
1.	Installation & training to users will be done by supplier free of cost.
Guarantee and Maintenance	
1.	Guarantee: Two years on equipment from the date of installation.
2.	CMC: CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years.
3.	The service engineer should be based in Rajasthan.


22

Documentation	
1. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.	
Demonstration	
1. Demonstration of equipment is must for technical approval.	

2. The bid is re-scheduled as follows:

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 11.05.2016	up to 1.00 P.M. 11.05.2016	from 3.00 P.M. 11.05.2016	up to 11:00 A.M. 23.05.2016	up to 1.00 P.M. 23.05.2016	from 3.00 P.M. 23.05.2016

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur