

**RAJASTHAN MEDICAL SERVICES CORPORATION LTD.**

(A Government of Rajasthan undertaking)

GANDHI BLOCK, SWASTHYA BHAWAN, TILAK MARG, JAIPUR

PHONE NO: 141-2223887, Fax No. 0141-5111040, CIN:U24232RJ2011SGC035067

E-mail Id- edepmmrc-rj@nic.in

No. F-8() RMSC/EPM/M-4/15-16/NIB-152/ Sec.E/ 1936

Dated: 20/4/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject:- Amendment under Bid document for the procurement Nd:Yag Laser and Phacoemulsification Machine under NIB No. F-8(RMSE/EPM/M-4/15-16/NIB-152/Sec.E/1003 Dated: 02.03.2016 and Addendum/Corrigendum no.1699 dated: 01.04.2016.

1. In reference to above cited subject and NIB No.-152, The Technical Specification of items (Section-V, Schedule of Supply, Point no. 3) are here by amended as below:-

Technical Specifications of Nd:Yag Laser and Phacoemulsification Machine

| S. No. | Description |
|--------|---|
| | <p style="text-align: center;">(A) <u>Technical Specifications for Nd:Yag Laser</u></p> <ol style="list-style-type: none"> 1. Mounted on imported slit lamp with original good quality motorized table. Slit lamp five step magnification 2. Touch screen control panel 3. Firing both from slit lamp joystic and footswitch 4. Solid state, 1064 nm wavelength, pulse length of 4 ns or less 5. Burst mode (1 to 3 bursts selectable) 6. Energy delivery upto 10 mJ in single burst mode 7. 8-15 microns spot size 8. Posterior offset at least 100 microns 9. Cone angle 16 to 20 degrees 10. Aiming red diode beam : 2-4 spot focusing with variable intensity 11. Capsulotomy and iridotomy lenses (two each) 12. Standard 220-240 V with 50 Hz power supply 13. Compatible UPS 14. Spare bulbs (12) 15. The product should have safety certificate from a competent authority CE / FDA (US). Copy of the certificate / test report shall be submitted in technical bid. 16. Guarantee shall be three years from the date of successful installation at consignee place. 17. CMC will be given @ 4 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years. |

97

| S. No. | <u>Technical Specification of Phacoemulsification Machine</u> |
|--------|--|
| | <p>Phacoemulsification unit with the following feature</p> <ol style="list-style-type: none"> 1. Phaco console <ol style="list-style-type: none"> a. Microprocessor controlled b. Phaco, I/A Vitrectomy, Diathermy/Bipolar Cautery function c. Suitable Surge control System d. Peristaltic pump design e. Memory settings for multisurgeon use f. Aspiration flow rate control 0-50 cc/mm g. Vacuum range 0 to at least 500 mmHg h. US frequency at least 25 KHz i. US power 0-100 % both panel and surgeon controlled j. Programmable pulse mode phaco with Variable pulse frequency and duration. k. Burst mode phaco l. Upgradable software 2. Stand <ol style="list-style-type: none"> a. infusion pole may be manual b. Trolley for keeping instruments on adjustable tray. 3. phaco handpiece <ol style="list-style-type: none"> a. At least 4 crystal titanium hand piece- two in number b. phaco needles (titanium) : 30 degrees (30 each with each unit) as specified below 4. I/A Handpiece <ol style="list-style-type: none"> a. one coaxial and one bimanual handpiece with each unit b. detachable tips compatible with disposable sleeves (two straight and two curved) with the coaxial I/A handpiece. 5. foot pedal with programmable detents 6. Vitrectomy : standard vitrectomy with at least 500 cuts per min and reusable Pneumatic/Electric Cutter. 7. Diathermy: standard bipolar Cautery with variable power adjustment 8. Accessories and disposables <ol style="list-style-type: none"> a. Tubings: <ol style="list-style-type: none"> 1. If reusable, at least 25 per unit 2. If single use, company should supply free of cost for at least three years (@ 275 per year) b. Phaco needles : <ol style="list-style-type: none"> 1. Straight 30 degrees 20 G - five 2. Curved 30 degrees 20 G – fifteen 3. Micro needle for 2.2 mm incision – ten c. Sleeve sets : Compatible with above needles – 200 units and for I/A tips- 200 units 9. Standard 220 volt, 50 Hz electrical supply with compatible online UPS with unit. 10. The machine should have safety certificate CE with notified Body/ FDA (US). Copy of the certificate shall be produced along with the technical bid. 12. Guarantee shall be three years from the date of successful installation at consignee place. 13. CMC will be given @ 4 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years. |

2. E-tenders are invited as per following time schedule:-

| Last Date for Sale of Bid Form | Last Date of Receipt of Bid Form | Date of Opening of Technical Bid |
|--------------------------------|----------------------------------|----------------------------------|
| 09.05.2016 Up to 11.00 A.M. | 09.05.2016 Up to 1.00 P.M. | 09.05.2016 From 3.00 P.M. |

Note: This corrigendum/addendum should be signed and annexed with bid document.

All other terms and conditions of bid shall remain the same.
This bears an approval of M.D., RMSCL, Jaipur.



Executive Director (EPM)
RMSCL, Jaipur
Dated: 20/4/16

No. F-8() RMSC/EPM/M-4/15-16/NIB-152// Sec.E/ 1936
Copy to:-

1. M.D.RISL, Yojna Bhawan, Jaipur
2. PA to M.D., RMSC, Jaipur
3. AGM (IT), RMSC, Jaipur
4. Guard File



Executive Director (EPM)
RMSCL, Jaipur

