

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmssc.nic.in

No. F-8() RMSC/EPM/M-2/2016-17/NIB-160/ 2629

Dated: 2/4/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical Specifications of items and revised bid submission date under NIB No. F-8()RMSC/EPM/M-2/2016-17/NIB-160/1691 dated 01.04.16

In Reference to subject cited above and NIB-160, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

1. Amended Technical Specifications of item Fetal Doppler:-

Use	
1.	Clinical Purpose To noninvasively detect fetal heart beats from the surface of the pregnant women's abdomen.
Technical Characteristics	
2.	Technical Characteristics Water proof probe of 2MHz frequency.
	Should have LCD / LCD / Seven Segment Display
	Auto Shut Off Facility to save Battery Power, Built in Speaker.
	Volume Control Facility and audio output for ear phone.
	Heart Rate Range should be from 50 to 210 bpm with accuracy of $\pm 2\%$.
	Should be water proof body.
Physical Characteristics	
3.	Physical Characteristics Should be handy, portable & easy to carry.
Power Requirements	
4.	Power Requirements Should have rechargeable battery with battery charger / adaptor.
Standards and Safety	
5.	Certificates Product should be European CE or USFDA certified.
Guarantee and Maintenance	
6.	Guarantee and Maintenance Three years on equipment.
	CMC: CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.
Documentation	
7.	Operating / User Manuals Operating / User manual with trouble-shooting guidelines should be provided with equipment.
	The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
Accessories & Consumables	
8.	Accessories & Consumables Should be supplied with Ultrasound gel bottles (10 No.)
	Should be supplied with good quality carrying case.
Demonstration	
9.	Demonstration Final technical approval after demonstration on patient.

2. Amended Technical Specifications of item Suction Pump - Foot Operated:-

Use	
1.	Clinical Purpose To aspirate fluids, secretions, or other foreign materials from a patient's airway by means of suction.
Technical & Physical Characteristics	
2.	Technical & Physical Characteristics Max. Vacuum / LPM: 0 - 600 mm Hg ± 10 mm regulable, flutter free vacuum control knob.
	Pump Type: Piston Pump
	Vacuum Gauge : 5 cm dia 0-760 mm Hg
	Dimensions approx: 32 X 17 X 30 cms

		Weight 2.5 kg max
		Foot pedal : Non corrosive metallic
		Transparent polycarbonate collection container capacity, approx: 1 L or more
		All parts can be autoclaved at 121°C.
Accessories & Consumables		
3.	Accessories & Consumables	Collection bottles, a vacuum gauge Microbial filter, tubing length minimum 150cm with combination of suction tip & angled connector.
Standards and Safety		
4.	Certifications	Product should be CE certified / ISI Marked
Guarantee		
5.	Guarantee	Three years on equipment.
Documentation		
6.	Operating / User Manuals	Operating / User manual with trouble-shooting guidelines should be provided with equipment. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
Demonstration		
7.	Demonstration	Final technical approval after demonstration on patient.

3. The bid is re-scheduled as follows:

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 03.06.2016	up to 1.00 P.M. 03.06.2016	from 3.00 P.M. 03.06.2016	up to 11:00 A.M. 21.06.2016	up to 1.00 P.M. 21.06.2016	from 3.00 P.M. 21.06.2016

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

Handwritten signature

Handwritten signature
Executive Director (EPM)
RMSCL, Jaipur

Handwritten mark