

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmssc.nic.in

No. F-8() RMSC/EPM/M-2/2016-17/NIB-161/ 2C28

Dated: 2/6/15

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical Specifications of items and revised bid submission date under NIB No. F-8()RMSC/EPM/M-2/2016-17/NIB-161/1692 dated 01.04.16

In Reference to subject cited above and NIB-161, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

1. Amended Technical Specifications of item CTG Machine (Cardio tocography) - Fetal Monitor:-

1. Fetal monitor for recording and analyzing the Fetal Heart Rate (FHR) on beat-to-beat basis.
2. Toco and Maternally sensed fetal movements. both manually and automatically detected
3. Should have facility of twin monitoring.
4. Should have built-in printer with multiple printing speed and automatic paper feeding facility.
5. Fetal monitor should have inbuilt rechargeable battery of **atleast 60 min. backup.**
6. Display of FHR upto Twins FHR1 & FHR2 & TOCO on **8" - 12" display.**
7. Heart rate offset mode to visually distinguish between the heart rates of twins with ability to offset the secondary FHR by atleast 20 BPM.
8. Uterine contractions alarm. Alarm delay facility, so that alarm is available only if the alarming condition is persistent for preset time. User friendly alarms and patient data.
9. Actual FHR in BPM.
10. Blinking corresponding to each Beat.
11. Patient Name / ID, date and time, Memory Backup/Graphical or Tabular trend for minimum **24 hours** with fast printing facility.
12. In-built / **Separate Acoustic** Stimulator with a separate marker on the graph for acoustic stimulators.
13. Ultrasound transducer should be multi crystal wide beam pulsed Doppler with frequency of 1MHz.
14. Fetal Heart Rate - measurements Range: 50-220 BPM.
15. External toco transducer which should be a sealed waterproof unit. Guard ring designed to reduce maternal respiration artifact. Measurement Range: 0-100 Units.
16. Event Market - Hand held, patient operated as well as front panel operated. Voltage 230V A.C. $\pm 10\%$, 50 Hz.
17. Product should be European CE / USFDA certified.
18. **Guarantee:** Three Years on equipment from the date of installation.
19. **CMC:** CMC will be given @ 5 % (of net rate- inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of three years.
20. Installation will be done by supplier free of cost.
21. List of spare parts and accessories to be supplied.
22. Company should have Service Network in Rajasthan.
23. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the technical bid along with compliance sheet as per technical specifications.
24. Should be supplied with following accessories:-

a) FHR Probe	- 2 Nos.
b) TOCO Probe	- 1 No.
c) Event Marker	- 1 No.
d) Fetal Simulator	- 1 No.
e) Probe Belt	- 3 Nos.
f) Ultrasound gel	- 10 Nos.
g) Compatible Paper roll	- 10 Nos.
h) Power Cable	- 1 No.
i) Movable Trolley with cover	- 1 No.
25. Rate of printing paper (Compatible with machine) should be provided by bidder in BOQ. Final rate of printing paper will be freezed after negotiation.
26. Final technical approval after demonstration on patient.

2. The bid is re-scheduled as follows:

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 03.06.2016	up to 1.00 P.M. 03.06.2016	from 3.00 P.M. 03.06.2016	up to 11:00 A.M. 21.06.2016	up to 1.00 P.M. 21.06.2016	from 3.00 P.M. 21.06.2016

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

Handwritten signature

Handwritten signature
Executive Director (EPM)
RMSCL, Jaipur

Handwritten signature