

**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmssc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: <http://rmssc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-2/17-18/NIB-263/ 91

Dated: 9/11/18

**CLARIFICATION/CORRIGENDUM/ADDENDUM****Sub:- Revised Technical Specifications of item *Cardiac Monitor with Defibrillator* under NIB No. F-8() RMSC/EPM/M-2/17-18/ NIB-263/5489 dated 10.11.17**


In Reference to subject cited above and NIB-263, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

**A. Amended Technical Specifications of item *Cardiac Monitor with Defibrillator***

1. Defibrillator should be Bi-Phasic.
2. Defibrillator should be ready for AED facility and non invasive pacing.
3. Should have a high resolution colour TFT display of minimum 6 inch or more.
4. Should have direct trim knob and direct function keys for mute and freeze.
5. Should have energy levels for defibrillation up to 200 joules or more.
6. Should be mains and battery operated. Internal battery should provide backup operation up to 2 -3 hours in monitoring or at least 90 defibrillation shocks should be delivered from fully charged battery.
7. Integrated external re-usable adult and paediatric paddles for defibrillation.
8. Should have non synchronised and synchronised cardio version.
9. Facilities of ECG pickup from paddles in case of ECG electrodes are not connected to the Defibrillator.
10. The charging time should be less than 7 seconds for charging up to 200 joules.
11. Should have integrated Printer.
12. Should be European CE/USFDA approved.
13. **Guarantee:** Three years on equipment from the date of installation and 06 months on accessories.
14. **CMC:** CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
15. Equipment should be supplied with compatible 3/5 lead ECG cable (2 Nos.) & Operating / user Manual (1 No.)
16. **Rates of Consumable:** Rates of consumable like Battery, 3/5 lead ECG cable, Disposable AED pad, thermal paper pack should be provided by bidder in BOQ in e-proc. Portal.
17. Installation will be done by supplier free of cost.
18. Firm should have service engineer network in Rajasthan.
19. Demonstration of equipment is must for technical approval.
20. The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.

**The bidders should provide Rates of Consumables in sealed envelope along with submission of bid security fees, tender fees and RISL Fees.**

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

  
Executive Director (EPM)  
RMSCL, Jaipur