



Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail : edepmrmisc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmssc.health.rajasthan.gov.in

No. F-8(322) RMSC/EPM/M-3/2017-18/NIB 285/ 873

Date: 26/2/18

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: -Revised Technical Specifications & Date Extension for NIB No. F-8(322) RMSC/EPM/M-3/2017-18/NIB-285/399 dated: 30.1.2018 for item Oxygen Concentrator.

In Reference to subject cited above and NIB-285, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

1. Revised Technical Specifications of Oxygen Concentrator

1. Oxygen concentrator to provide oxygen from ambient air.
2. Oxygen concentration measured at the flow meter by oxygen sensing device (OSD).
3. *Sound level should be < 50dB.*
4. Superior grade of molecular sieve.
5. Maintenance free rotary proppet valve.
6. Oxygen purity should be 92% ± 2%.
7. Oxygen output, approx: 0 - 5 LPM.
8. *Pressure range should be 6 - 8 PSI.*
9. Single outlet to maintain output presser & flow/ Double outlet or flow splitter for oxygen Delivery.
10. Oxygen tube of 2m length must be provided with.
11. With two humidifier bottles and two cabinet filters.
12. Power requirements: 220 V/50 Hz.
13. Device is produced by ISO 9001 certified manufacturer.
14. Device should be certified ISO 80601-2-69:2014.
15. Bottles should be made of autoclavable polycarbonate.
16. With bilateral oxygen nozzle.
17. Oxygen failure or Low Oxygen Alarm facility should be available for below 80% oxygen.
18. Device is safety certified according EUROPEAN CE / USFDA.
19. **Guarantee:** Two years on equipment from the date of installation.
20. **CMC:** CMC will be given @ 5 % exclusive GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of two years.
21. **Supplied with:-**
 - 1) 1 x spare set of tubing.
 - 2) Two spare sets of humidifier bottles.
 - 3) Two spare sets of cabinet filters.
 - 4) User manual with trouble shooting guidance, in English.
 - 5) Technical manual with maintenance and first line technical intervention instructions, in English.


- 22) Rates of accessories like Reusable oxygen tubing, Oxygen Mask for Adult, Pediatric & neonatal should be provided by bidder in BOQ in e-proc. Portal.
- 23) The service engineer should be based in Rajasthan.
- 24) Demonstration of equipment is must for technical approval.
- 25) Installation will be done by supplier free of cost.
- 26) The company should mention the make & model of the quoted item and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

2. Date extension of Bid:-

Extended dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
13.3.2018 11:00 AM	13.3.2018 06:00 PM	14.3.2018 11:00 AM

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same. Decision of purchase committee dated 23.2.2018 headed by M.D. RMSCL


Executive Director (EPM)
RMSCL, Jaipur