

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmse-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmse.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-2/17-18/NIB-292/ 583

Dated: 7/2/18

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical Specifications Revised bid submission date of item Emergency First Aid Kit with Respirator Care Equipment under NIB No.F-8()RMSC /EPM/M-2/17-18/NIB-292/6053 dated 28.12.17

In Reference to subject cited above and NIB-292, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

A. Amended Technical Specifications of item Emergency First Aid Kit with Respirator Care Equipment:-

Kit with Ventilator:-

Specifications

Technical Specifications of Emergency First Aid Kit with Respirator Care Equipment

Each Kit Shall Contain Automatic & Manual Resuscitator, Oxygen Inhalation Suction Pump Intubation Set, Diagnostic Instrument, Needle Holder & Infusion Dressings etc. Automatic Resuscitator – One Set

(a) Ventilator (Automatic Resuscitator)

It shall have electrically / battery/pneumatically operated Ventilator to meet all Emergency Situation for Respiratory Support. Minimum battery backup of 4 Hrs. or more (if electrically / battery operated). The Equipment shall be easy to Operate with Non-invasive & Invasive ventilation of Adult & children, automatically with wide range of tidal volume from 50-1500ml, Should have a wide frequency range 8to 50bpm, Modes of ventilation:- Volume controlled, pressure support , synchronized intermittent mandatory ventilation (SIMV), Assist /control mode , CPAP ,Inbuilt PEEP , Alarm required : minute volume , pressure , PEEP apnoea , occlusion , high respiration rate , disconnection , low battery alarm(if electrically / battery operated)- Weight not more than 6 Kg,Should have Selectable Fio2% :21% to 100% .

Product should be European CE certified / USFDA Certified.

Following accessories:-

1. Three sizes face masks with tubing
2. Oxygen cylinder 3 ltr. Water capacity bull nose, Light Aluminium.
3. Single stage gauge regulator

A. MANUAL RESUCITATOR - ONE SET

ISI marked manually operated resuscitator (Material-Silicon) autoclavable with double inlet valves for air and oxygen. Attachment: Face masks 3 Nos. (1, 3, & 5) should also be given.

C. OXYGEN INHALATION - ONE SET

Oxygen flow control from 0 to 10 lit/min simply turning by 0 knob. Supplied with nasal oxygen catheter, Adult & child poly mask & Ventimask (Adult-1 No) for concentration of air & oxygen with tubing.

D. SUCTION PUMP - ONE SET

Manually operated suction pump to aspirate the mucus, blood or other with.

1. Performance: maximum suction pressure of 38 mm hg
2. Peak Flow rate : 70 l/minute
3. Portable & with a mounting bracket.
4. Collection bottle 1000 ml capacity and sterilizable
5. Overflow protection insured.
6. It should be supplied with one set all standard accessories like Microbial filter, tubing with combination of suction tip & angled connector.
7. Product should be CE certified / ISI Marked.

E. INTUBATIONS SET - ONE SET

LED Laryngoscope with three blades and handle, endotracheal tube with cuff & plain all sizes(3.5 to 8.5), reusable type, endotracheal connection set of 12, Magill's forceps. Introducing forceps, mouth bite .

F. DIAGNOSTIC INSTRUMENTS - ONE SET

- 1) ISI mark Stethoscope - One No
- 2) ISI mark aneroid sphygmomanometer - One No.
- 3) Percussion hammer - One no
- 4) Tongue spatula - One no.
- 5) Examination torch with two dry cell plastic body - 1 no.
- ISI mark clinical thermometer - 1 no

G. INFUSIONS, DRESSINGS & OTHER - ONE SET

- 1) I.V. rod in two sections -1 no.,
- 2) I.V. set disposable - 1 no.
- 3) Adhesive plaster USP size 5cm. x 5mt. - One roll.
- 4) Sterilized gauge 20cm. x 20cm. - 2 Nos.
- 5) Rolled bandage 4 nos.
- 6) Dressing scissor SS 15cm. S/B - one no.,
- 7) Dressing forceps SS 12.5cm - One.
- 8) Tissue forceps SS - one no.
- 9) Haemostatic forceps straight SS size 12.5cm. one no.,
- 10)
- 11) Needle holder SS 15cm. - One no.
- 12) Disposable syringes 5ml. - 2 nos.
- 13) Disposable needle 22/24g. (1 pack of 10-12)
- 14) Splints - one set (Pack of 6).
- 15) Cylinder key.

General Terms & Conditions

- A. Weight should be approximately 20 ± 5 kg. And size should be either 24" or lesser. The company should be ISO certified.
 - B. **Consumable items:** Face mask, endotracheal tube, Breathing circuit, disposable syringes & disposable needles.
 - C. **Guarantee:** Three year from the date of supply / delivery.
 - D. The service engineer should be based in Rajasthan.
 - E. Demonstration of Equipment is must for technical approval.
- The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

Kit Without Ventilator

Specifications

Technical Specifications of *Emergency First Aid Kit with Respirator Care Equipment*

Each Kit Shall Contain Manual Resuscitator, Oxygen Inhalation Suction Pump Intubation Set, Diagnostic Instrument, Needle Holder & Infusion Dressings etc.

A. MANUAL RESUCITATOR - ONE SET

ISI marked manually operated resuscitator (Material-Silicon) autoclavable with double inlet valves for air and oxygen. Attachment: Face masks 3 Nos. (1, 3, & 5) should also be given with :-

1. Three sizes face masks with tubing
2. Oxygen cylinder 3 ltr. Water capacity bull nose, Light Aluminium.
3. Single stage gauge regulator

B. OXYGEN INHALATION - ONE SET

Oxygen flow control from 0 to 10 lt./min simply turning by 0 knob. Supplied with nasal oxygen catheter, Adult & child poly mask & Ventimask (Adult-1 No) for concentration of air & oxygen with tubing.

C. SUCTION PUMP - ONE SET

Manually operated suction pump to aspirate the mucus, blood or other with.

1. Performance: maximum suction pressure of 38 mm hg
2. Peak Flow rate : 70 l/minute
3. Portable & with a mounting bracket.
4. Collection bottle 1000 ml capacity and sterilizable
5. Overflow protection insured.
6. It should be supplied with one set all standard accessories like Microbial filter, tubing with combination of suction tip & angled connector.
7. Product should be CE certified / ISI Marked.

D. INTUBATIONS SET - ONE SET

LED Laryngoscope with three blades and handle, endotracheal tube with cuff & plain all sizes(3.5 to 8.5), reusable type, endotracheal connection set of 12, Magill's forceps. Introducing forceps, mouth bite .

E. DIAGNOSTIC INSTRUMENTS - ONE SET

- 1) ISI mark Stethoscope - One No
- 2) ISI mark aneroid sphygmomanometer - One No.
- 3) Percussion hammer - One no
- 4) Tongue spatula - One no.
- 5) Examination torch with two dry cell plastic body - 1 no.
- ISI mark clinical thermometer - 1 no

F. INFUSIONS, DRESSINGS & OTHER - ONE SET

- 1) I.V. rod in two sections -1 no.,
- 2) I.V. set disposable - 1 no.
- 3) Adhesive plaster USP size 5cm. x 5mt. - One roll.
- 4) Sterilized gauge 20cm. x 20cm. - 2 Nos.
- 5) Rolled bandage 4 nos.
- 6) Dressing scissor SS 15cm. S/B - one no.,
- 7) Dressing forceps SS 12.5cm - One.
- 8) Tissue forceps SS - one no.
- 9) Haemostatic forceps straight SS size 12.5cm. one no.,
- 10) Needle holder SS 15cm. - One no.

Handwritten signatures and initials at the bottom right of the page.

- 11) Disposable syringes 5ml. – 2 nos.
 12) Disposable needle 22/24g. (1 pack of 10-12)
 13) Splints - one set (Pack of 6).
 14) Cylinder key.

General Terms & Conditions

- A. Weight should be approximately 20 ± 5 kg. And size should be either 24" or lesser. The company should be ISO certified.
 B. **Consumable items:** Face mask, endotracheal tube, Breathing circuit, disposable syringes & disposable needles.
 C. **Guarantee:** Three year from the date of supply / delivery.
 D. The service engineer should be based in Rajasthan.
 E. Demonstration of Equipment is must for technical approval.
 The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

Note:- Firm has to submit rate of quoted item in sealed envelope with ventilator and without ventilator.

B. The bid is re-scheduled as follows:

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
<u>08.02.2018</u> <u>3:00 PM</u>	<u>09.02.2018</u> <u>1:00 PM</u>	<u>09.02.2018</u> <u>3:00 PM</u>	<u>26.02.2018</u> <u>11:00 AM</u>	<u>26.02.2018</u> <u>6:00 PM</u>	<u>27.02.2018</u> <u>3:00 PM</u>

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.


 7.2.2018
 Executive Director (EPM)
 RMSCL, Jaipur