

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011SGC035067

E-mail : edepmrmssc-rj@nic.in
Website: www.rmssc.nic.in

No. F-8() RMSC/EPM/M-2/2016-17/NIB-196/ 5021

Dated: 28/11/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical Specifications and revised bid submission date for item LED Phototherapy Unit under NIB No. F-8()RMSC/EPM/M-2/2016-17/NIB-196/4820 dated 11.11.16 due on dated 02.12.16.

In Reference to subject cited above and NIB-196, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

1. Amended Technical Specifications of item LED Phototherapy Unit:-

1. Lifetime of LED 20,000 to 30,000 hours
2. Spectral Irradiance from 30 to 60 $\mu\text{W}\cdot\text{cm}^{-2}\cdot\text{nm}^{-1}$ at 45 cm distance between bed and light
3. Should have multilevel intensity control from minimum to maximum
4. At the tilted position, the irradiance is 30 $\mu\text{W}\cdot\text{cm}^{-2}\cdot\text{nm}^{-1}$ at 45 cm distance between bed and light
5. Wavelength is 450 – 480-nm, and it's free from UV and IR radiation.
6. Effective surface area - 200 X 400 mm within a irradiance ratio of 0.4 (min/max irradiance)
7. Digital (LCD) Timer for monitoring therapy hours (resettable) & lamp usage hours (non resettable)
8. Have feather touch key pad
9. Cooling Fan/ heat sink is provided to dissipate the heat created by LED's
10. Light head should be compact to use along with the Radiant warmer & provided with tilting facility so that the unit is not coming directly under warmer.
11. Smooth Height adjustment mechanism & seamless height adjustment from minimum to maximum height
12. Minimum height – 1150 mm or less from the floor to use near the mother bed
13. Maximum height – 1500 mm or more from the floor to use with the incubator and Radiant warmer.
14. Coating: Epoxy/powder coated body for scratch and rust prevention and
15. PU (Poly Urethane)/ABS/Fiber coating for plastic
16. Mobility: Three/four castors; two rear castors provided with brakes
17. The base of the unit is such that it will go beneath any Incubator/bed/trolley, with minimum of 100 mm floor clearance
18. Should be supplied with white LED for patient examination
19. The unit should be supplied with reusable eye masks, Quantity: 10
20. The product should be CE marked and the manufacturer should be ISO: 13485 certified (Certificate to be submitted)
21. The product should have regulatory approval IEC 60601-1(certification to be submitted)
22. **Two years guarantee after installation**
23. **CMC: CMC will be given @ 5 % (of net rate- inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of two years.**

2. The bid is re-scheduled as follows:

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 02.12.2016	up to 1:00 P.M. 02.12.2016	from 3:00 P.M. 02.12.2016	up to 11:00 A.M. 06.12.2016	up to 1:00 P.M. 06.12.2016	from 3:00 P.M. 06.12.2016

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur