



**RAJASTHAN MEDICAL SERVICES CORPORATION LTD.**

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - [edepmrmsc-rj@nic.in](mailto:edepmrmsc-rj@nic.in)

No. F-8( ) RMSC/EPM/M-1/17-18/NIB-391/ 2018/ 22)

Dated: 23-1-19

**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Subject:** - Revised Bid Schedule of item Deep Freezer(-20°C) and Amended Technical Specifications of Bid document of item, under No. F-8( ) RMSC/EPM/M-1/18-19/NIB-391/4571 date: 24.12.18

In reference to subject cited above and NIB-391, the representations received from the various firms and issues raised by the Bidders, are examined by the competent Authorities and technical committee and approved by purchase committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document Amended Technical Specification of items as below:-

**Technical Specifications of Item - Deep Freezer(-20°C)**

**Deep Freezer(-20°C)**

1. The system should be vertical standing model.
2. The system should have 7 segment/LCD electronic control unit with digital temperatures display on panel.
3. The outer body of system is made of rust & chip resistant Mild steel and inner body high grade stainless steel/ABS plastic.
4. The system should have volume capacity of 350 liters ± 10 liters
5. The temperatures range should be (-20°C to -25°C) or below.
6. The system should be microprocessor controlled.
7. Uninterrupted 24 X 7 electronic temperature circular chart recorder / integrated digital temperature cum controller with inbuilt 7 days inkless circular chart recorder with supply of 200 free charts and should be provision to connect with central (temperature) monitoring system/data logger for monitoring the temperature.
8. The system should have CFC Free refrigerant.
9. The entire cabinet should be able to maintain uniform temperature in each drawer.
10. The system should have PUF insulation.
11. The system should have minimum 4 to 7 storage drawer/shelves with single/individual lids.
12. The system should have a built- in audible & visible alarm, if temperature fluctuates from its set-point range & door open warning system.
13. Firm should bring digital temp. monitor/thermometer (NABL Accredited) to check entire cabinet temp at the time of demonstration.
14. The system should be CE certified.
15. A servo stabilizer inbuilt/external to be supplied with the system
16. The system should be operable at 220-230 V & 50-60 Hz
17. Supply with four wheel base for easy operation with locking mechanism.
18. Guarantee: Three years on equipment.
19. CMC: CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for Five years (on yearly basis) after completion of Guarantee period of three years.
20. Each Firm/ Company is requested to submit Brochure of the equipment with detailed specification Pre- installation requirement, installation and training to be done by firm on

*[Handwritten signature]*



site Local service engineer should be available in 24-48 hr time Support to be provided for application on phone User list and satisfactory certificate to be provided from Indian institution.

**Revised bid schedule:-**

E-bids are invited as per following revised time schedule:-

Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
<u>05.02.2019</u> <u>11:00 AM</u>	<u>05.02.2019</u> <u>6:00 PM</u>	<u>06.02.2019</u> <u>11:00 AM</u>

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website [www.rmssc.nic.in](http://www.rmssc.nic.in) and [sppp.raj.nic.in](http://sppp.raj.nic.in). Please note that individual bidder will not be intimated."

**Note:-**

Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.

  
Executive Director (EPM)  
RMSCL, Jaipur