

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmesc.nic.in

No. F-8 (51)RMSC/EPM/M-5/17-18/NIB-232/ 2402

Dated: 5/6/17

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended technical Specifications, date extension and other terms and conditions of Bid document for the ENT Equipments under Short-Term NIB No. F-8 (51)RMSC/EPM/M-5/17-18/NIB-232/2035 Dated 05.05.2017.

In Reference to above cited subject and NIB-232, the technical specification of items are here by amended as below:-

1. Revised Technical Specifications of Digital Audiometer -

Essential:

Power supply: 220 Volts, 50Hz

Transducer: headset with earphones (TDH 39) & supraaural ear cushions (M X41 AR), insert earphones (ER-3A), bone vibrator (Radio ear b-71), Talk over microphone, Talk back microphone & loudspeaker (Qty-2)

Patch cords

Stimuli Type: Pure tones (continuous, pulse tone, interruptor) & Warble tones (continuous, pulse tone, interruptor), Narrowband noise, speech noise, speech (monitored live voice and recorded voice)

Frequencies: 250Hz to 8000Hz in octaves for air conduction 250Hz to 4000Hz in octaves for bone conduction

Intensity:

-10 dB HL to 120dB HL in steps of 5dB for air conduction

-10 dB HL to 70 dB HL in steps of 5 dB for bone conduction from 250Hz to 4000Hz

Power supply: UPS/Battery backup & mains.

Self-calibration

- Two years Guaranty to be given with the equipment.
- Minimum of two routine maintenance services/inspections to be provided annually.
- CMC will be given @5% (of net rate-inclusive of excise duty and exclusive of VAT/CST etc.) plus service tax (as applicable) & yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion guarantee period of two year.
- The Manufacture should have ISO 13485.
- Should have safety certificate from a competent authority or valid detail electrical and functional safety test. Copy of the certificate/test report shall be produced along with the technical bid.
- Final technical approval after demonstration.
- Price quoted should be inclusive of their installation at site.

2. Revised Technical Specifications of OAE Analyzer

Essential

1. **Power Supply:** 220 Volts, 50Hz
2. **Type of OAE:** TEOAEs, DPOAEs and screening ABR capable.
3. Frequency range of TEOAE should be 4000 Hz or above and for DPOAE should be 5000 Hz or above.
4. Stimuli intensity level up to 80 dB SPL for TEOAEs and up to 70 db SPL or above for DPOAEs.
5. Automatic F1 and F2 ratio calculation.
6. Low battery warning
7. Serviceable probes; probe mount
8. Cleaning kit
9. Printer with minimum 10 paper roll.
10. Power supply: Rechargeable Battery operated
11. Can be used as stand alone or with a PC
12. Self-calibration
13. Separate ear tip for infants and adults (50 no. adult & 50 no. infant/pediatric)
14. Spare probe head/coupler 3 Nos. should be supplied free of cost with each machine.
15. Data Analysis should be fully automated (It should display cross spectrum, DP gram, Pass/Reffer test report).
16. The equipment should be CE approved / USFDA.



17. Should have safety certificate from a competent authority or valid detail of electrical and functional safety test. Copy of the certificate/test report shall be produced along with the technical bid.
18. Two years Guaranty to be given with the equipment.
19. Minimum of two routine maintenance services / inspections to be provided annually.
20. CMC will be given @5% (of net rate-inclusive of excise duty and exclusive of VAT/CST etc.) plus service tax (as applicable) & yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion guarantee period of two year.
21. Price quoted should be inclusive of their installation at site.
22. Final technical approval after demonstration.

3. Revised Technical Specifications of Sound Proof Room - No Change

In Reference to above cited subject and NIB-232, the last date for sale of bid, receipt of bid and date of opening of technical bid is hereby extended as below:-

Existing Date			Extended date		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
08.06.2017 11:00 AM	08.06.2017 01:00 PM	08.06.2017 03:00 PM	21.06.2017 11:00 AM	21.06.2017 01:00 PM	21.06.2017 03:00 PM

Note: Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms and conditions of bid shall remain the same.

This bears the approval of M.D., RMSCL, Jaipur

Executive Director (EPM)
RMSCL, Jaipur