

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011SGC035067E-mail : edepmrmisc-rj@nic.in
Website: www.rmsc.nic.in

No. F-8() RMSC/EPM/M-2/16-17/NIB-194/ 4986

Dated: 17/11/16

CLARIFICATION/CORRIGENDUM/ADDENDUM**Sub:- Revised Technical Specifications of item New Born Weighing Scale-Digital under NIB No. F-8() RMSC/EPM/M-2/16-17/NIB-194/4745 dated 07.11.16 due on dated 24.11.16.**

In Reference to subject cited above and NIB-194, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

1. Amended Technical Specifications of item New Born Weighing Scale-Digital:-

1. Electronic scale for weighing new born babies
2. Measuring range 0 to approx 15 kg
3. Minimum graduation: 5 g
4. With tare function
5. On switch and auto-off
6. Auto-calibration with each switch-on
7. Large LED display readable in low light working situations, display cover durable plastic
8. Display in kg and lbs, easy switch between kg and lbs
9. Reading time max 5 seconds
10. Zero weighing adjustment
11. Freeze reading feature
12. Smooth surface/finishing allows for easy cleaning/disinfection.
13. All vital parts made of rust proof materials
14. Horizontal levelling with height adjustable feet
15. Splash proof and shock resistant light-weight body
16. Power requirements: 220 V/50 Hz with 30 minutes battery backup.
17. Power consumption: 150 W or less.
18. Device is produced by ISO 9001 certified manufacturer.
19. Device should be CE certified/ ISI certified.
20. Guarantee: One years on equipment from the date of installation.
21. Machine should be supplied with verified calibration certificate along with stamp by weights & Measures department.

Supplied with:

22. 1 x spare set of fuses
23. User manual with trouble shooting guidance, in English/Hindi.
24. Technical manual with maintenance and first line technical intervention instructions, in English.
25. Training and installation at end-user site

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

**Executive Director (EPM)
RMSCL, Jaipur**