

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065
CIN:U24232RJ2011SGC035067

E-Mail : edepmrmisc-ri@nic.in
Website : www.rmssc.nic.in

No. F-8(261) RMSC/EPM/M-3/17-18/NIB-235/ 4070

Date: 29/9/17

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: -Revised Technical Specifications and other terms and conditions of Bid document for VENTILATORS F-8 (261) RMSC/EPM/M-3/17-18/NIB-235/3434 dated: 17.8.2017.

In Reference to subject cited above and NIB-235, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document & Technical Specification of items (Section-V; Schedule of Supply, Point no. 3) as below:-

Technical Specifications for Basic Ventilator

For Basic Ventilator, point no. 11 & 12 should be read as:-

Point no. 11:

"Should have inbuilt / integrated Air source of same make" instead of 'Should have integrated Air source of same make'.

Point no. 12:

"It should be compatible with central gas supply of air & O₂ in ICU setup, wherever applicable" instead of 'It should be compatible with central gas supply of air & O₂ in ICU setup'.

Technical Specifications for Advanced Ventilator

For Advanced Ventilator, point no. 2 should be read as:-

Point no. 2:


"The unit should be external compressor based of same make for precise gas delivery" instead of 'The unit should be compressor based of same make for precise gas delivery.

The last date for sale of bid, receipt of bid and date of opening of technical bid hereby extended as below:-

Extended dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
16.10.2017 11:00 AM	16.10.2017 01:00 PM	16.10.2017 03:00 PM

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.
This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur