

मुख्यमंत्रा निःशुल्क जाय योज  
**RAJASTHAN MEDICAL SERVICES CORPORATION LTD.**

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN:U24232RJ2011SGC035087 E-Mail -rpsc@nic.in, edepmrpsc-rj@nic.in

F-8(297) RMSC/EPM/M-3/17-18/NIB-257/646 Date: 9/2/18

**OFFICE ORDER**

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of F-8(297) RMSC/EPM/M-3/2017-18/NIB-257/5281 Dated: 27.10.2017, technical bid opening date 21.12.2017. The list of bidders declared responsive/non responsive as per decision taken in the purchase committee dated 9.2.2018 headed by M.D., RMSCL.

S. No.	Name of Tenderer Firm	Name of Quoted Item	Responsive/ Non responsive	Remark
1.	M/s ALM ElectroMed Systems & Services, 488/5/C-3, First Floor, Block 'A' Market, Dilshad Garden, Delhi-110095 Email:-info@alm-electromed.in	BIPAP MASK VENTILATOR WITH WELLCUSHIONED FACE MASK WITH HARNESS	Non responsive	Technically not approved (AVAPS or Equivalent mode not available)
2.	M/s ResMed India, Pvt.Ltd., 205,2nd Floor, Southern park, D-2, District Centre, Saket, New Delhi-110 017 Email:-vinay.rattan@resmed.co.in, enquiries@resmed.co.in	BIPAP MASK VENTILATOR WITH WELLCUSHIONED FACE MASK WITH HARNESS	Responsive	Responsive
3.	M/s APRAMEYA Engineering H-4/24, Akruiti Apartment, Shastri Nagar, Naranpura, Ahmedabad-380063 E-mail: aprameyaengg@gmail.com	BIPAP MASK VENTILATOR WITH WELLCUSHIONED FACE MASK WITH HARNESS	Responsive	Responsive

Online financial bids of responsive bidders for above item will be opened on dated 12.2.2018 at 11.00 A.M. Demonstration report can be obtained from office during office time.

  
Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd.  
Jaipur