

**Rajasthan Medical Services Corporation Limited, Jaipur**

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: [rmssc@nic.in](mailto:rmssc@nic.in)

CIN : U24232RJ2011SGC035067

Website : [www.rmssc.nic.in](http://www.rmssc.nic.in)

No.: F-8(109) RMSC/EPM/M-1/B.A. 113.4/NIB-226/2017/ 1121

Dated: 23/3/17

**-: OFFICE ORDER :-**

List of Bidders declared Responsive/Non Responsive as per Documentary & Technical Evaluation of NIB No. F-8(109) RMSC/EPM/M-1/B.A. 113.4/NIB-226/2017/ 641 Dated: 17.02.17 Technical bid opening date 06.03.17 The list of bidders declared Responsive/Non Responsive is as follows:-

Sr. No	Name of Item	Name of Firm	Responsive / Non-Responsive as per Demonstration Reports and Documentary Evaluation	Remarks
01	02	03	04	05
1.	Horizontal Laminar Air Flow	M/s Authentic Instruments 31, RIICO Industrial Area Jhotwara, Kamani Choraha Jaipur- 302012	Responsive	Nil
		M/s Associated Scientific Technologies C_A & 9, Ambey Garden Libaspur, Samaypur Ext. Industrial Area, Delhi	Non-Responsive	Rejected as per Demonstration Report due to:- <ul style="list-style-type: none"> <li>• Work table size was 4' x 2' x 2' instead of 4' x 2' x 3.5'</li> <li>• No provision of hanging the blood bag was provided in cabin</li> </ul>
		M/s Sudarshan Chemicals (Regd), Sudarshan Building, Jhalawar Road Kota	Non-Responsive	Rejected as per Demonstration Report due to:- <ul style="list-style-type: none"> <li>• Only on and off timer for UV light and no total time usages of UV light</li> <li>• Air Flow of Laminar is 2' feet high instead of 3.5' feet</li> </ul>

\* Detail Demo report can be obtained from office, during office hours

Online financial bids of Responsive bidder for above item will be opened on dated 27.03.2017 at 11:00 A.M.

  
**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation,**  
**Jaipur**