

**Rajasthan Medical Services Corporation Limited, Jaipur****Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005**

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011ISGC035067

Website: http://rmisc.health.rajasthan.gov.in

No. F-8( ) RMSC/EPM/M-1/17-18/NIB-261/2018/ 780

Dated: 20/2/18.


**OFFICE ORDER**

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of NIB No. F-8( ) RMSC/EPM/M-1/17-18/NIB-261/5459 dated 08.11.17, technical bid opening date 03.01.18. The list of bidders declared responsive/ non responsive as per decision taken in the purchase committee dated 20.02.2018 headed by M.D., RMSCL:-

S. No.	Name of Firm	Quoted items	Responsive /Non responsive	Remarks
1.	M/s Meril Diagnostics Pvt. Ltd., Survey No. 135/2/B , Bilakhia House Muktanand Marg, Chala, Gujrat	Elisa Reader and Washer	Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically recommended किया गया।
2.	M/s Transasia Bio-Medicals Ltd. Transasia House, 8 Chandivali Studio Road Andhari East, Mumbai	Elisa Reader and Washer	Non-Responsive	<ul style="list-style-type: none"> <li>• तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically Not approved (standard accessories not provided ) किया गया।</li> <li>• Complete Elisa reader with computer, printer and UPS not brought during demo.</li> <li>• Inbuilt printer unable to printout – 12 rows+8columns reading of elisa plate.</li> <li>• Software was not able to interpret to show result on screen correctly.</li> <li>• Washer – Only 8 way wash head was demonstrated not 12 way wash head.</li> </ul>
3.	M/s Sun Bio System , 250, 2nd Floor, Vardhaman Plaza, Vasundhara Enclave , Delhi	Elisa Reader and Washer	Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically recommended किया गया।
4.	M/s Sudarshan Chemicals (Regd) ,Sudarshan Building, Jhalawar Road Kota	CENTRIFUGE MACHINE- (16 TUBES)	Non-Responsive	<ul style="list-style-type: none"> <li>• तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically Not approved (standard accessories not provided ) किया गया।</li> <li>• Centrifuge machine was with bush motor. Machine was not with SS Tube Carrier.</li> </ul>

5.	M/s Remi Elektrotechnik Ltd., Remi House, 11 CAMA INDL Estate, Goregaon (E), Mumbai	CENTRIFUGE MACHINE- (16 TUBES)	Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically recommended किया गया।
6.	M/s Remi Elektrotechnik Ltd., Remi House, 11 CAMA INDL Estate, Goregaon (E), Mumbai	CENTRIFUGE MACHINE - (36 TUBES)	Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically recommended किया गया।

Online financial bids of responsive bidders for above item will be opened on dated 21.02.18 at 03:00 P.M. Demonstration report can be obtained from office during office time.

  
20.2.2018  
Executive Director (EPM)  
RMSCL, Jaipur