

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065
CIN : U24232RJ2011SGC035067

E-mail: edepmrmse-rj@nic.in
Website : www.rmse.nic.in

No. F-8(28) RMSC/EPM/M-5/2016-17/NIB-212/ 467


Dated: 3/2/17

OFFICE ORDER

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of No. F-8(28) RMSC/EPM/M-5/2016-17/NIB-212/6382 Dated 29.12.2016, technical bid opening date 17.01.2017. The list of bidders declared responsive/ non responsive as per decision taken in the purchase committee dated 03.02.2017 headed by M.D., RMSCL as follows:-

| S no | Name of Firm | Name of Item | Responsive/Non-Responsive | Remarks |
|------|---|---|---------------------------|---|
| 1. | M/s Kamal Coach Works Pvt. Ltd., Sp-282, RIICO Industrial Area, Phase-II, Newai, Dist Tonk Email: info@kamalcoach.com | Equipping and Fabrication of ALS Ambulances | Responsive | NIL |
| 2. | M/s Instromedix (India) Pvt. Ltd., A5-A6, Pragati Chambers, Ranjit Nagar Commercial Complex, New Delhi-110008 Email: instro@del2.vsnl.net.in | Equipping and Fabrication of ALS Ambulances | Non-Responsive | Technical Bid Non-Responsive due to non-fulfilment of technical specification criteria w.r.t following point:- 1. Multipara Monitor quoted by bidder does not have European CE Certificate as specified in Technical Specification 2. External defibrillator quoted by bidder is a semi-automatic External defibrillator instead of Fully Automated External defibrillator 3. Electric Suction Machine quoted by bidder has maximum continuous working time of 30 min as per catalogue submitted by bidder instead of required 90 min. |

Online financial bids of responsive bidders for above item will be opened on dated 08.02.2017 at 11:00 A.M.


Executive Director (EPM)
Rajasthan Medical Services Corporation,
Jaipur