



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN:U24232RJ2011SGC035067 E-Mail -rmsc@nic.in, edepmrmisc-ri@nic.in

F-8(86)RMSC/EPM/M-4/17-18/NIB-277/ 963 Date: 28/12/18

OFFICE ORDER

List of bidders declared responsive/non responsive for Item **Slit Lamp** as per documentary evaluation & technical evaluation of F-8(86)RMSC/EPM/M-4/2017-18/NIB-277/5496 Dated: 13.11.2017, technical bid opening date 28.12.2017. The list of bidders declared responsive/non responsive for Item **Slit Lamp** as per decision taken in the purchase committee dated 28.2.2018 headed by M.D., RMSCL.

S. No.	Name of Tenderer Firm	Responsive/ Non responsive	Remark
1.	M/s Towa Sales Corporation, 223 OKHLA Industries Estate, New Delhi 110020	Responsive	Responsive
2.	M/s Healthskape Medicals Pvt. Ltd. 4-425 A, Sharam Industrial Estate Walldhat Road, Goregaon(East) Mumbai-400063	Non Responsive	Non Responsive RMSCL के स्पेसिफिकेशन में Point no. 2 पर It Should have 5 step magnification with 10 x eye Pieces है जबकि फर्म द्वारा Five Steps magnification with 12.5x eyepieces दिखाया गया है। तथा आरएमएससीएल स्पेसिफिकेशन Point no. 3 पर It should have Continuously variable Slit Width 0 to 14 mm है। फर्म द्वारा Slit width 0 to 12 mm दिखाई गई।
3.	M/s Mehra Eyetech Private Limited, NO 22, Main Road, Sakthi Nagar, Puducherry-605013-	Non Responsive	Non Responsive M/s Mehra Eyetech ने Technical Compliance sheet एवं Original Product catalog में Five Steps magnification with 12.5x eyepieces का टेक्स्टर प्रेषित किया गया है। जबकि निगम द्वारा Five Steps magnification with 10x eyepieces चाहा गया है। M/s Mehra Eyetech की निविदा को विभागीय तकनीकी स्पेसिफिकेशन के अनुरूप नहीं पाये जाने के कारण Non-responsive किया जाता है। Not eligible for demonstration.
4.	M/s Aurolab ,Madurai-	Non Responsive	Non Responsive Mandatory document EM-II नहीं लगाया गया है। अतः Non-responsive किया जाता है। It is mandatory document. Not eligible for demonstration).
5.	M/s Devine Meditech E-41/5 Okhla Industrial Area, Phase-II New Delhi-110020	Non Responsive	Non Responsive Mandatory document EM-II नहीं लगाया गया है। अतः Non-responsive किया जाता है। It is mandatory document. Not eligible for demonstration).

Online financial bids of responsive bidders for above item will be opened on dated 05.3.2018 at 11.00 A.M. Demonstration report can be obtained from office during office time.

28.2.2018
Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd.
Jaipur