

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN:U24232RJ2011SGC035067 E-Mail -rmssc@nic.in, edepmrmssc-ri@nic.in

F-8(261)RMSC/EPM/M-3/17-18/NIB-235/ 239 Date: 19/11/18

OFFICE ORDER

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of F-8(261) RMSC/EPM/M-3/2017-18/NIB-235/3434 Dated: 17.8.2017, technical bid opening date 16.10.2017. The list of bidders declared responsive/non responsive as per decision taken in the purchase committee dated 18.1.2018 headed by M.D., RMSCL.

S. No.	Name of Tenderer Firm	Name of Quoted Item	Responsive/ Non responsive	Remark
1.	M/s Aprameya Engineering H-4/24, Akruiti Apartment, Shastri Nagar, Naranpura, Ahmedabad-380063 Email:- aprameyaengg@gmail.com	Ventilators (Basic and Advanced Ventilator)	Responsive	NIL
2.	M/s Schiller Healthcare India Pvt. Ltd., Advance House, Makwana Road, Off Andheri Kurla Road, Marol Naka, Andheri (E), Mumbai- 400059 Email:- mail@schillerindia.com	Ventilators (Basic and Advanced Ventilator)	Responsive	NIL
3.	M/s Rohanika Electronics & Medical Systems C-49, Okhla Phase-I, New Delhi-110020 Email : devdutt@rohanika.com	Ventilators (Basic and Advanced Ventilator)	Non responsive	Technically Non Responsive For Advanced Ventilator Firm Point no. 2 compressor is of EKOM company. OEM certificate was produced but USFDA/European CE certification was for EKOM compressor only and not for evolution 3E. Hence compressor cannot be considered as same make, so technically not approved. For Basic Ventilator not approved (PRVC mode quoted as optional advance ventilation mode is not equivalent as per specifications.)
4.	M/s Surbhi Meditek, 138, First Floor, Metropolis Tower, Nirmohi Nagar, Near Purani Chungi, Ajmer Road, Jaipur-302019 Email:-surbhi.meditek@gmail.com	Ventilators (Basic and Advanced Ventilator)	Responsive	Responsive for item Advanced Ventilator (For Advanced Ventilator Firm has minor deviations from technical specifications. However these deviations will not effect the functioning of machine, hence technically approved.) Non Responsive for Basic Ventilator (For Basic Ventilator not approved (PRVC & PV tool are quoted as optional advanced ventilation mode which are not equivalent as per specifications.)
5.	M/s Cardiolabs Healthcare India Pvt. Ltd., Plot no. 1512, JLPL Industrial Park, Sector 82, Mohali-140 306, Punjab Email:-info@cardiolabshealthcare.com	Ventilators (Basic and Advanced Ventilator)	Non responsive	Not eligible for demonstration (CE document submitted by Firm was not European CE/USFDA which is a mandatory document)
6.	M/s Medtronic India Medtronic Pvt. Ltd., 4th Floor, Tower A & B SAS Tower Medanta The Medicity Complex, Sector 38, Gurgaon-122001 Haryana, India Email:-dharampal.rawat@medtronic.com	Ventilators (Basic and Advanced Ventilator)	Non responsive	Not eligible for demonstration (BF-XVIII not uploaded, It is mandatory document)

Online financial bids of responsive bidders for above item will be opened on dated 22.1.2018 at 11.00 A.M. Demonstration report can be obtained from office during office time.

KM
19.11.2018
Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.