

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmsc-rj@nic.in

Website:www.rmssc.nic.in

No F-8(40) RMSC/EPM/M-5/16-17/NIB-222/ 1435

Dated: 31/3/17

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended technical Specifications, date extension and other terms and conditions of Bid document for (1) Abbes / Butyro Refractometer (2) Conductivity Meter + pH Meter (3) Kjeldal Digestion Unit with Kjeldhal Distillation Unit under NIB No. F-8(40) RMSC/EPM/M-5/2016-17/NIB-222/644 Dated 17.02.2017.

In Reference to above cited subject and NIB-222, the technical specification of items are here by amended as below:-

Revised Specification of Abbes / Butyro Refractometer

1. The Refractometer should be digital, compact in design & with reliable optics.
2. RI measurement range: below 1.28 to 1.72 or more
3. BRIX measurement range: 0 to 100%
4. Temperature control: Automatic with built in Peltier Thermo Module.
5. Temperature control range with Peltier : 10°C to 85°C.
6. Resolution (RI): up to 5th decimal or more
7. Accuracy (RI): up to 4th decimal or more
8. Light source: LED lamp with average lifetime of 100,000 h
9. Data Interfaces: RS 232 port for printer connection, USB port.
10. Recording of data on system's internal memory.
11. Compliance: Inbuilt 21 CFR part 11
12. Automatic temperature correction.
13. The provision for setting the limit (fail & pass) for the respective test sample should be inbuilt.
14. Should be supplied with standard solution for calibration of RI & BRIX.
15. Manufacturer of Equipment should be ISO 9001 certified.
16. Equipment should be US FDA / European CE certified.
17. **Guarantee: Two years** on equipment, accessories and all peripherals / sub systems from the date of installation, which Includes 2 visits and unlimited breakdown calls by service/application support engineer during the Guarantee period.
18. **CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.**
19. Installation will be done by supplier free of cost including all parts like wires, tubes, joints & attachments, small fixtures etc.

Revised Technical Specification of Conductivity Meter + pH Meter -No Change

Revised Technical Specification of Kjeldal Digestion Unit with Kjeldhal Distillation Unit -No Change

In Reference to above cited subject and NIB-222, the last date for sale of bid, receipt of bid and date of opening of technical bid is hereby extended as below:-

Existing Date			Extended date		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
06.04.2017 11:00 AM	06.04.2017 01:00 PM	06.04.2017 03:00 PM	18.04.2017 11:00 AM	18.04.2017 01:00 PM	18.04.2017 03:00 PM

Note: Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms and conditions of bid shall remain the same.

This bears the approval of M.D., RMSCL, Jaipur


Executive Director (EPM)
RMSCL, Jaipur