

**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-ri@nic.in

Website:www.rmssc.nic.in

No F-8(40) RMSC/EPM/M-5/16-17/NIB-223/ 1434

Dated: 31/3/17

**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Subject: Amended technical Specifications, date extension and other terms and conditions of Bid document for (1) Electronic Weighing Balance (Electronic Type – Digital) (2) Vacuum Oven with Vacuum Pump (3) Open Top balance (Electronic Type – Digital) under NIB No. F-8(40) RMSC/EPM/M-5/2016-17/NIB-223/645 Dated 17.02.2017.**

In Reference to above cited subject and NIB-223, the technical specification of items are here by amended as below:-

**Revised Specification of Open Top balance (electronic type – Digital)****Open Top balance (electronic type – Digital)**

1. **Weigh accurately least count= 10mg**, EMFC technology, Operating temperature 10 to 45° C, RS 232 to USB converter.
2. Fully automatic time and temp. Controlled internal Calibration and balance should be capable to adjust itself.
3. Auto zero setting.
4. Weighing capacity up to 3 KG.
5. Balance should have- liquid crystal display (LCD) with multi parameter display illuminated by white back light, stainless steel weighing pan.

**Terms & Conditions:**

1. Equipment should be **ISI/CE certified**.
2. Equipment model should be approved from department of legal metrology.
3. **Guarantee:** Two years on equipment from the date of installation.
4. **CMC:** CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after **Guarantee period of two years**.
5. Installation & Training will be done by supplier free of cost.
6. The service engineer should be based in Rajasthan.
7. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

**Revised Specification of Vacuum Oven with Vacuum Pump****Vacuum Oven with Vacuum Pump**

1. Should be Double Walled (Size 12"x12"x16")
2. Outer Chamber made of M.S. Powder Coated
3. Inner chamber is made of STAINLESS STEEL (304 grade).
4. Thermal loss is prevented by filling mineral wool (Insulation) in space between the outer body & Inner chamber.
5. Provided with Door Gasket for proper sealing.
6. Fitted with heavy hinges & door closing device.
7. See through transparent window of toughened glass is provided at the front.
8. Adjustable two air controlling valves are fitted for creation of vacuum in the chamber & release.
9. Heating elements are made of high quality Nichrome/Kanthal wire are embedded around the chamber.
10. Temperature is controlled by electronic digital temperature indicator cum controller from **below 50°C to 130°C or more ± 0.1 °C**. fitted on the front panel for easy operations.
11. Temperature is displayed by Digital LED with SV and PV display at front of the oven.
12. Front Panel is provided with ON/OFF switch, temperature controller and indicators
13. Supplied with cord and plug.
14. Special Specification :

Temperature Controller : By Microprocessor Based PID Digital Temperature Indicator cum Controller

Temperature Sensitivity : 0.1 °C

Relay : Solid State Electronic relay with protective heat sink

15. should be sullied with Vacuum Pump of following specification :
- It should be oil free diaphragm vacuum pump
  - Ultimate Vacuum should be 2 metric cube per hour.
  - Suction/ Flow capacity should be 7mbar or below 7mbar.
  - It should be less power consumption below 200 Watt.
  - Standard warranty is 3 years or more as given by manufacturer.

**Other Terms & Conditions:**

- Manufacturer of Equipment should be ISO 9001 certified.
- Equipment should be CE certified.
- Guarantee: Two years** on equipment, accessories and all peripherals / sub systems from the date of installation, which Includes 2 visits and unlimited breakdown calls by service/application support engineer during the Guarantee period.
- Installation will be done by supplier free of cost including all parts like wires, tubes, joints & attachments, small fixtures etc.
- Firm should mention all the pre-installation requirements in technical bid.**
- CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.**

**Revised Technical Specification of Electronic Weighing Balance (Electronic Type – Digital)-No Change**


In Reference to above cited subject and NIB-223, the last date for sale of bid, receipt of bid and date of opening of technical bid is hereby extended as below:-

Existing Date			Extended date		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
06.04.2017 11:00 AM	06.04.2017 01:00 PM	06.04.2017 03:00 PM	18.04.2017 11:00 AM	18.04.2017 01:00 PM	18.04.2017 03:00 PM

**Note: Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.**

All other terms and conditions of bid shall remain the same.

This bears the approval of M.D., RMSCL, Jaipur

  
Executive Director (EPM)  
RMSCL, Jaipur