

Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
 CIN:U24232RJ2011SGC035067

E-mail : edepmrmrsc-rj@nic.in
 Website:www.rmssc.nic.in

No F-8(40) RMSC/EPM/M-5/16-17/NIB-225/ 143C

Dated: 31/3/17

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended technical Specifications, date extension and other terms and conditions of Bid document for (1) Heating Mantles - 500 mL , 2 Litre (each) (2) Soxhlet Extraction Unit (3) Nephelometer / Turbidity meter (4) Quartz Water Distillation Unit (5) Pensky Marten Flash Point Test Apparatus (6) Polarimeter with Sodium Lamp assembly (Indigenous) (7) U. V. Cabinet with Long & Short wave UV Lamp under NIB No. F-8(40) RMSC/EPM/M-5/2016-17/NIB-225/647 Dated 17.02.2017.

In Reference to above cited subject and NIB-225, the technical specification of items are here by amended as below:-

Revised Specification of Heating Mantles - 500 mL , 2 Litre (each)

- Should be durable, chemical resistant polypropylene housing/non corrosive aluminum housing .
- Flexible heating element ensures minimum risk of flask breakage.
- These controlled and grounded electro mantles feature a built-in power regulator and safety earth (ground) screen. Unique ventilation slots or without ventilation slots keep the exterior safe to the touch even while in use.
- The replaceable coiled heating element is suspended within a thermal insulating cartridge to provide maximum heat transfer and support.
- Ceramic fiber insulation keeps the chemical-resistant polypropylene case cool.
- Built-in power controller has dial markings from 1 to 10 for repeatability.
- Mantles should have indicator lights for "power on" and "heater on" .
- All mantles are double fused and include a single support rod clamp that accepts a 1/2" rod.
- The stainless steel grounding screen provides added protection against shock in the event of spillage or flask breakage.
- Mantles include 2-meter power cord with India plug .
- Should be supplied 2 No's extra heating element with each Heating Mantels & distillation unit of borosil make 500mL , 2L (each).
- Equipment should be CE Certified .
- Guarantee:** Two years on equipment from the date of installation on all Parts of equipment including metal fitting / tubes / Glass parts etc. against manufacturing defects.
- CMC:** CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.

Revised Technical Specification of Soxhlet Extraction Unit -No Change

Revised Technical Specification of Nephelometer / Turbidity meter -No Change

Revised Technical Specification of Quartz Water Distillation Unit -No Change

Revised Technical Specification of Pensky Marten Flash Point Test Apparatus -No Change

Revised Technical Specification of Polarimeter with Sodium Lamp assembly (Indigenous) -No Change

Revised Technical Specification of U. V. Cabinet with Long & Short wave UV Lamp -No Change


In Reference to above cited subject and NIB-225, the last date for sale of bid, receipt of bid and date of opening of technical bid is hereby extended as below:-

Existing Date			Extended date		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
11.04.2017 11:00 AM	11.04.2017 01:00 PM	11.04.2017 03:00 PM	25.04.2017 11:00 AM	25.04.2017 01:00 PM	25.04.2017 03:00 PM

Note: Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms and conditions of bid shall remain the same.

This bears the approval of M.D., RMSCL, Jaipur


Executive Director (EPM)
RMSCL, Jaipur