



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur – 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011SGC035067

E-mail : edepmrmse-rj@nic.in
Website: www.rmse.nic.in

No. F-8(213) RMSC/EPM/M-3/16-17/NIB 217/

862

Dated: 7-3-17

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical Specifications and Bid Extended date of OFFICE FURNITURE/HOSPITAL FURNITURE & OTHER SUPPLIES under NIB No. F-8(213) RMSC/ EPM/M-3/16-17/NIB-217/623 dated 16.2.2017

In Reference to subject cited above and NIB-217, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in B

id document & Technical Specification of items as below:-

Amended Technical Specifications HOSPITAL FURNITURE & OTHER SUPPLIES

S. No.	Description
	<p style="text-align: center;"><u>Technical Specification of items</u></p> <p>1. iv Stand:</p> <ol style="list-style-type: none">1. Stainless Steel (S.S.) Tubular pipe frame with S.S. nut.2. Heavy duty S.S. base on five castors (50mm) for stability (Non rusting).3. 2 hooks provided made up of SS rod for I.V. fluids.4. Thickness of S.S. Tubular pipe 16 Gauge of 1.25" diameter.5. S.S. adjustable rod from 135 to 240 cms. <p>General Terms & Conditions:</p> <ol style="list-style-type: none">1. Guaranty: Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.2. Packing shall be of good quality to avoid damage or scratches during transit.3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.5. The RMSCL may do pre dispatch inspection at supplier store/premises/factory. <p>2. Patient Attendant Benches with mattress (Attendant Bed for Cottage & Cubicles) :</p> <ol style="list-style-type: none">1. Attendant bed should be Box type with top opening with lid.2. Frame work made of Square / Tubular pipe frame M.S. Tube 16 Gauge of 1.25".3. M.S. frame work mounted on PVC stumps.4. 5 cm foam padded top on CRCA Sheet 20 Gauge thick or more, covered with thick rexine

[Signature]

- (rexine mounted on cloth), density of foam should not be less than 40.
5. Pre-treated and epoxy powder coated.
 6. Overall approx. size: 180 L x 60 W x 46H cms, (tolerance $\pm 5\%$)

General Terms & Conditions:

1. **Guaranty:** Guaranty of one year on against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises /factory.

3. Patient Stretcher Trolleys:

1. Trolley made of pipe frame M.S. Tube 16 Gauge 1" round for horizontal & 1.25" for vertical.
2. Standard length & width and height 210 (L) X 56(W) X 65-95 (H) cm
3. Should have four noiseless 15 cm diameter heavy duty castor wheels, 2 Nos. with locks for break movement.
4. Wheels should not be welded out the trolley as repair / replacement of wheels later on is possible.
5. The trolley should have facility to keep Oxygen cylinder.
6. Should have weight bearing capacity of around at 200 Kg (Patient + O₂ cylinder).
7. Head-end of the stretcher top size around 2.25 - 2.5 feet should be adjustable to put at deferent angles i.e. 15 to 75 degree.
8. Detachable top should have this adjustable head end.
9. Arrangement for I/V stand and for supporting O₂ (oxygen) cylinder.
10. 1.5" mattress at top, density of foam should be ≥ 40 .
11. Pre-treated and epoxy powder coated finished.

General Terms & Conditions:

1. **Guaranty:** Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.

4. Bed side Screen (3 Panel):

1. Made of M.S. Tubular frame in 3 folds.
2. Should be mounted on 5 cm antistatic castors with curtains wire on hooks.
3. Size: 66" H x 96" L (tolerance $\pm 5\%$)
4. Thickness of pipe 16 Gauge of 0.75" diameter.
5. Pre-treated and epoxy powder coated.
6. The screen cloth should be green or blue colour and washable.

General Terms & Conditions:

1. **Guaranty:** Guaranty of one year against any manufacturing defect from the date of supply /

Handwritten signature

delivery / installation.

2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.

5. "Foot Step(Single Step)":

1. Made of 16 Gauge 1" M.S. tubular frames fitted with PVC stump.
2. Top covered with M.S. sheet of 16 Gauge.
3. Overall approx size : 18"(L) X 9"(W) X 9" (H), (tolerance+5%).
4. Pre-treated and epoxy powder coated.

General Terms & Conditions:

1. **Guaranty:** Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.

6. "Revolving Stool" :

1. M.S. tubular four legs (1.25"OD x 16 Gauge thick tube) fitted with PVC stumps.
2. Foot ring support of stainless steel.
3. 18" to 27" adjustable height S.S. top (18 gauge thickness) through screw system & it should be stable at max. height.
4. Pre-treated and epoxy powder coated.

General Terms & Conditions:

1. **Guaranty:** Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.

The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.

7. "Medicine Cabinet (Standard Size)"

1. Overall approx size: 76 (L) X 46(W) X 81 (H) cm. (Tolerance $\pm 5\%$)
2. Frame work of MS tubes of 16G mounted on 10 cm antistatic castors.
3. Two S.S. shelves of 20G with three side railing on top shelf of suitable height.
4. Two drawer under the each shelf
5. Pre-treated and epoxy powder coated.

General Terms & Conditions:

1. **Guaranty:** Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.

JS

2. Packing shall be of good quality to avoid damage or scratches during transit.
 3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
 4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
- The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.

8. Crash Cart for ICU:

1. Six removable bins & S.S. Frame, two polystyrene storage unit with three drawers, each can be locked.
2. Fitted with 125 mm castors (two with break) complete with corner buffers.
3. Oxygen Cylinder cage/holder, Electric Lamp.
4. I.V Rod, Cardiac Massage Board & three laminated shelves (SS)
5. Pre-treated & epoxy powder coated.
6. Overall approx size: 940 mm (L) X 490 mm (W) X 153 mm (H) - (Tolerance $\pm 5\%$) made of 20G sheet and 16G Pipe of dia. 1".

General Terms & Conditions:

1. **Guaranty:** Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.

In Reference to above cited subject and NIB-218, the last date for sale of bid, receipt of bid and date of opening of technical bid hereby extended as below:-

Extended dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
28-03-2017	28-03-2017	28-03-2017
11:00 AM	01:00 PM	03:00 PM

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur

