

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrm-sc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rm-sc.nic.in

No. F-8()RMSC/EPM/M-2/2016-17/NIB-187/ 136

Dated: 11/1/17

OFFICE ORDER

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of NIB No. F-8()RMSC/EPM/M-2/2016-17/NIB-187/4438 dated 21.10.16, technical bid opening date 05.12.2016. The list of bidders declared responsive/ non responsive as per decision taken in the purchase committee dated 10.01.17 headed by M.D., RMSCL:-

| S. No. | Name of Firm | Quoted items | Responsive /Non responsive | Remarks |
|--------|--|--|----------------------------|---|
| 1. | M/s Garg Scientific & Surgical Co., Ajmer | CTG Machine (Cardio tocography)- Fetal Monitor | Responsive | -- |
| 2. | M/s Medical Care System, Jaipur | CTG Machine (Cardio tocography)- Fetal Monitor | Non Responsive | Firm did not submit past performance as per bid conditions |
| 3. | M/s BPL Medical Technologies Private Limited, Jaipur | CTG Machine (Cardio tocography)- Fetal Monitor | Non Responsive | European CE Certificate submitted in bid for the quoted item was not valid |
| 4. | M/s Maestros Electronics & Telecommunications Systems Limited, Navi Mumbai | CTG Machine (Cardio tocography)- Fetal Monitor | Non Responsive | Demonstrated item was technically not recommended by the technical committee. |

Online financial bids of responsive bidders for above item will be opened on dated 13.01.2017 at 03.15 P.M.


Executive Director (EPM)
RMSCL, Jaipur