

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur
Phone no. 0141-2223887, Fax no. 0141-2228065
CIN : U24232RJ2011SGC035067

E-mail: edepmrmsc-rj@nic.in
Website : www.rmssc.nic.in

No. F-8(40) RMSC/EPM/M-5/16-17/NIB-222/ 2502

Dated: 9/6/17

OFFICE ORDER

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of No. F-8(40) RMSC/EPM/M-5/16-17/NIB-222/644 Dated 17.02.17, technical bid opening date 18.04.17. The list of bidders declared responsive/ non responsive as per decision taken in the purchase committee dated 08.06.17 headed by M.D., RMSCL as follows:-

S no	Name of Firm	Name of Item	Responsive/ Non-Responsive	Remarks
1.	M/s Agaram Industries, 73 Nelson Road, Aminikikarai, Chennai	Abbes / Butyro Refractometer	Non-Responsive	ISO9001 and EUCE Certificate (Mandatory Document) Not submitted in bid.
2.	M/s Anton Paar India Pvt, Ltd, 582 Phase V, Udhog vihar Industiral area, Gurgaon.	Abbes / Butyro Refractometer	Responsive	NIL
3.	M/s Atago Instruments Pvt Ltd, Unit 103, Western Edge II, Opp Megathane Bus depot, Off Western Express highway, Bornali (east), Mumbai	Abbes / Butyro Refractometer	Non-Responsive	In Demonstration Technically Not Approved

Encloser - Demo Report Enclosed.

Online financial bids of responsive bidders for above item will be opened on dated 13.06.2017 at 11:00 A.M.

202
Executive Director (EPM)
Rajasthan Medical Services Corporation,
Jaipur

[Signature]

[Signature]

Demo date: 26/05/2017

Sr. No.	Specifications	Model & Make	M/s. Anton Paar India Pvt. Ltd., Gurgaon	M/s ATAGO India, Mumbai
1	The Refractometer should be digital, compact in design & with reliable optics.		Yes	Yes
2	RI measurement range: below 1.28 to 1.72 or more		Yes	Not comply RI range 1.32-1.70
3	BRIX measurement range: 0 to 100%		Yes	Yes
4	Temperature control: Automatic with built in Peltier Thermo Module.		Yes	Yes
5	Temperature control range with Peltier : 10°C to 85°C.		Yes	Not comply 5°C.-75 °C.
6	Resolution (RI): up to 5th decimal or more		Yes	Yes
7	Accuracy (RI): up to 4th decimal or more		Yes	Yes
8	Light source: LED lamp with average lifetime of 100,000 h		Yes	Yes
9	Data Interfaces: RS 232 port for printer connection, USB port.		Yes	Yes
10	Recording of data on system's internal memory.		Yes	Yes
11	Compliance: Inbuilt 21 CFR part 11		Yes	Not comply , 21 CFR part 11 Via PC
12	Automatic temperature correction.		Yes	Yes
13	The provision for setting the limit (fail & pass) for the respective test sample should be inbuilt.		Yes	Yes
14	Should be supplied with standard solution for calibration of RI & BRIX.		Yes	Yes
15	Manufacturer of Equipment should be ISO 9001 certified.		Yes	Yes
16	Equipment should be US FDA / European CE certified.		Yes	Yes
17	Guarantee: Two years on equipment, accessories and all peripherals / sub systems from the date of installation, which Includes 2 visits and unlimited breakdown calls by service/application support engineer during the Guarantee period.		Yes	Yes
18	CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.		Yes	Yes
19	Installation will be done by supplier free of cost including all parts like wires, tubes, joints & attachments, small fixtures etc.		Yes	Yes
Remarks			Technically Recommended	Technically not recommended

Handwritten signature and date: 26/05/17

Signature of Chief Field Analyst, Rajasthan, JAIPUR

Handwritten signature and date: 26/05/17