

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmisc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-2/18-19/NIB-398/ 1389

Dated: 23/1/2019


OFFICE ORDER

The following bidders are declared responsive/non responsive on the basis of documentary & technical evaluation of proposal submitted in response to NIB No. F-8()RMSC/EPM/M-2/18-19/NIB-398/4612 dated 28.12.18, (technical bid opening date 06.02.2019) as per the decision taken in the purchase committee dated 09.04.19 headed by M.D., RMSCL:-

S. No.	Name of Firm	Quoted items	Remarks	Responsive /Non responsive
1	M/s Fanem Medical Devices India Pvt. Ltd., #279, 4th Main, 4th Phase, Peenya Industrial Area, Bangalore-560058 E-mail: info@fanem.in	LED Phototherapy Unit	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically approved किया गया।	Responsive
2	M/s Phoenix Medical Systems Pvt. Ltd., 32/4, Jawaharlal Mehru Salai, Ekkattuthangal, P.O. Box -3205 Chennai-600032 Email:sales@pmsind.com	LED Phototherapy Unit	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically approved किया गया।	Responsive
		Open Newborn Care System on Trolley (Radiant Warmer)	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically not approved किया गया। Reason: Specification Point no.17- it should have adjustable audio visual alarm facility for overheating – non adjustable alarm Point no. 31- radiant warmer shall be provided with attachment to hold for X-Ray film, report, stethoscope, CPAP/Ventilator tubing and hand sanitizer, also provision of B-type oxygen cylinder - no provision for CPAP/Ventilator tubing, oxygen cylinder holder should be in vertical position Point no.33- Should have standard IV pole (sturdy; non rusting; medical grade stainless steel; adjustable to max height of 6 feet from the ground level), monitor tray (12X10 inches; 270 deg swivel; axed at level of warmer display) and storage drawer – in demo sample IV pole stand is rusted	Non-Responsive
3	M/s Zeal Medical Pvt. Ltd., 4/19-A, Piramal Industrial Estate, S.V. Road, Goregaon (W), Mumbai-400062 Email: info@zealmedical.com	LED Phototherapy Unit	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically approved किया गया।	Responsive
		Open Newborn Care System on Trolley (Radiant Warmer)	तकनीकी मूल्यांकन किये जाने पर फर्म की निविदा आईटम के लिये दस्तावेजी रूप से अयोग्य पाई गई।	Non-Responsive

4	M/s Medi Waves INC, B-68, G.T. Karnal Road Industrial Area, Delhi-110033 Email: sales@medi- waves.com, anand_sachin@yahoo.com	LED Phototherapy Unit	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically not approved किया गया। Reason: specification point no. 18.- the base of the unit is such that it will go beneath any incubator/pad/trolley, with minimum of 100mm floor clearance- not available point no. 19-should be supplied with white LED for patient examination – not available	Non-Responsive
		Open Newborn Care System on Trolley (Radiant Warmer)	तकनीकी मूल्यांकन किये जाने पर फर्म की निविदा आईटम के लिये दस्तावेजी रूप से अयोग्य पाई गई।	Non-Responsive

Online financial bids of responsive bidders for above items will be opened on 24.04.19 at 03:00 p.m.



Executive Director (EPM)
RMSCL, Jaipur