

**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail : edepmmssc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website www.rmssc.health.rajasthan.gov.in

No. NIB-928/GEM/2026/B/7161034

584

Dated: 23/7/26

Clarification/Corrigendum/Addendum

Subject:- NIB No. GEM/2026/B/7161034 Dated: 30.01.2026 for Item "Transcutaneous Billirubinometer".

In Reference to subject cited above and GEM/2026/B/7161034 Dated: 30.01.2026 the following Corrigendum/Addendum is issued for inclusion in bid document as below:-

1. Date extension of Bid :-

Existing Dates		Extended Dates	
Last Date & Time of online Submission of bid	Date & Time of Online Opening of Technical Bid	Last Date & Time of online Submission of bid	Date & Time of Online Opening of Technical Bid
24.07.2026 15:00 P.M.	24.07.2026 15:30 P.M.	10.08.2026 15:00 P.M.	10.08.2026 15:30 P.M.

2. AMENDMENT IN TECHNICAL SPECIFICATION OF TRANSCUTANEOUS BILIRUBINOMETER & PAYMENT CONDITION.

SR.No	REVISED TECHNICAL SPECIFICATION OF TRANSCUTANEOUS BILIRUBINOMETER
1.	Intended Use For Extreme Preterm, Preterm Full term or Post Term Babies Up To 20 Post Natal Days
2.	It should measure reading at sternum and forehead or ear (Scaphoid Fossa)
3.	In Case of reading measures through ear then testing tip replace free of cost during guarantee period in case of any damage.
4.	SERUM BILIRUBIN RANGE - 0 TO 20mg/dl
5.	Accuracy +/- 1.5 mg/dl
6.	Repeatability +/- 2 mg/dl
7.	Weight ≤250gm (with battery)
8.	Light Source should have life of more than 100000 measurements.
9.	Battery Type - Lithium ion/Ni-MH
10.	Battery Life-Two Year Minimum
11.	Measurement With Fully Charge Battery - minimum 200 single measurements
12.	Power Supply (can be operated with AC Adaptor/Charger)

Signature valid

Digitally signed by A. Omprakash

Designation: Director

Date: 2026.07.23 17:44:59 IST

Reason: Approval

RajKaj Ref No.:
23465864

M e-Sign





	Input- 100-240 VAC, 50/60Hz, 0.4 A-2A
13.	Should have user friendly control panel.
14.	It should have facility to displayed observed bilirubin in mg/dl.
15.	It should supply with: Charger unit with a checker, AC adapter, Carrying case and wrist strap, power cable adapter set.
16.	- Quoted model should be USFDA approval/European CE (From Notified Body)/BIS. - Quoted model should have valid CDSCO Licence. Copy of licence should be submitted along with technical bid.
17.	No hidden costs of disposable /reusable tip require.
18.	Should have large easy to read display.
19.	WARRANTY- Warranty of 3 years from the date of installation.
20.	Documentation- User or Technical or Maintenance Manuals (In English) to be supplied.
21.	Other General Term and Conditions- - The company shall maintain the make and model name/number of the quoted equipment and submit the technical brochure of the quoted model in the technical bid along with compliance sheet as per technical specification. - Preventive maintenance schedule- As per OEM(Mention PM Schedule) - Calibration Schedule- As per OEM(Mention Schedule) - Life of equipment- As per OEM

New point added in ATC –

10. Part payment of equipment will be made as decided by MD, RMSCL. 30% payment released after warranty period expired.

Note :- All other terms & conditions remains the same.

Director (P)
RMSCL, Jaipur

Signature valid

Digitally signed by A. Omprakash
Designation: Director
Date: 2026.06.27 14:44:59 IST
Reason: Approved

RajKaj Ref No.:
23465864
M e-Sign