

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: <http://rmisc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-1/17-18/NIB-253/2018/ 957

Dated: 5/3/18

OFFICE ORDER


List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of NIB No. F-8() RMSC/EPM/M-1/17-18/NIB-253/2017/3831 dated 12.09.17, technical bid opening date 04.01.18. The list of bidders declared responsive/ non responsive as per decision taken in the purchase committee dated 28.02.2018 and 05.03.18 for item (1) Laparoscope Set For Training With CO2 Insufflator and (2) Ultrasonic Cutting And Coagulating Device For Open & Laparoscopic Surgery respectively headed by M.D., RMSCL:-

S. No.	Name of Firm	Quoted items	Responsive /Non responsive	Remarks
1.	M/s Divine Meditech Pvt. Ltd. 159, Railway Road, Nangal, Dist: Ropar, Punjab	Laparoscope Set For Training With CO2 Insufflator	Non-Responsive	<p>तकनीकी मूल्यांकन किये जाने पर फर्म की निविदा आईटम के लिये दस्तावेजी रूप से योग्य नहीं पाई गई। Doesn't provided shortcomings demanded from firm -</p> <ol style="list-style-type: none"> Notarized BF-XVIII. Attested copy of Average Annual Turnover by notary public. Notarized copy of BF-V on a stamp paper of Rs. 200 Statement of plant and machinery (BF-VIII) on a stamp paper of Rs. 50 Technical Compliance Sheet for quoted item along with original product catalogue. Clarification regarding authenticity of USFDA/European CE Certificate Clarification whether All the quoted items are manufactured by same firm
2.	M/s Borze Heager India Pvt. Ltd B-125 to 127, First Floor, Pacific Business Park, Sahibabad Industrial Area Delhi NCR- 201010	Laparoscope Set For Training With CO2 Insufflator	Non- Responsive	<ul style="list-style-type: none"> तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically Not approved किया गया <p>Camera head is devoid of any mark of company</p> <p>Resolution and progressive scan technology is not present in camera head.</p> <p>Plastic/metal tray from Oem for storage and sterilization not</p>

				present Extendible arm for monitor mount not present
3.	KARL STORZ ENDOSCOPY INDIA PVT LTD 11 floor, Dr. Gopal Das Bhawan 28, Barakhamba Road, New Delhi 110001	Laparoscope Set For Training With CO2 Insufflator	Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically recommended किया गया।
4.	Olympus Medical Systems India Pvt Ltd Ground Floor, Tower- C, SAS Tower, The medicity Complex Sector-38, Gurgaon, Haryana, India	Laparoscope Set For Training With CO2 Insufflator	Non-Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically Not approved किया गया LED Lamp life is of 10000 hours only instead of demanded 25000 hours Extendible arm for monitor mount not present
5.	STRYKER INDIA PRIVATE LTD 10 floor, Block Two, Sect. 49, Sohna Road Gurgaon 122002, Haryana, India	Laparoscope Set For Training With CO2 Insufflator	Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically recommended किया गया।
6.	Johnson and Johnson Limited LBS Marg, Mulund(W) Mumbai 400080	Ultrasonic Cutting And Coagulating Device For Open & Laparoscopic Surgery.	Responsive	क्रय समिति द्वारा विचार विमर्श उपरांत फर्म को Responsive माने जाने का निर्णय लिया गया
7.	Olympus Medical Systems India Pvt Ltd Ground Floor, Tower- C, SAS Tower, The medicity Complex Sector-38, Gurgaon, Haryana, India	Ultrasonic Cutting And Coagulating Device For Open & Laparoscopic Surgery.	Non-Responsive	As per specification "Should have facility of rotating hand switch adapter with integrated bilateral switches to enable precise operation of system by hand for hooks and blades" – Hook feature not available
8.	India Medtronic Pvt. Ltd 4 floor, Tower A&B SAS Tower	Ultrasonic Cutting And Coagulating Device For Open & Laparoscopic	Non-Responsive	As per specification "Should have facility of rotating hand switch adapter with integrated bilateral switches to enable

	Medanta The Medicity complex, Haryana	Surgery.		precise operation of system by hand for hooks and blades" -- Hook feature not available
--	---------------------------------------	----------	--	---

Online financial bids of responsive bidders for above item will be opened on dated 06.03.18 at 11:00 A.M. Demonstration report can be obtained from office during office time.


 5.3.2018
 Executive Director (EPM)
 RMSCL, Jaipur