

# Rajasthan Medical Services Corporation Ltd.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN:U24232RJ2011SGC035067 E-Mail -rmssc@nic.in, edepmrmssc-ri@nic.in

F-8(315)RMSC/EPM/M-3/2017-18/NIB-273/ 956

Date: 5/3/18


## Office Order

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of F-8(313)RMSC/EPM/M-3/2017-18/NIB-273/5503 Dated: 13.11.2017, technical bid opening date 10.01.2018. The list of bidders declared responsive/non responsive as per decision taken in the purchase committee dated 05.03.2018 headed by M.D., RMSCL.

S. No.	Name of Tenderer Firm	Name of Quoted Item	Responsive/ Non responsive	Remark
1.	M/s York Scientific Industries Pvt. Ltd., 8/6E, Site IV, Sahibabad Industrial Area, Distt. Ghaziabad (U.P.)-201 010	Deep Freezer – Horizontal	Non-Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा कोटेड आईटम डेमोस्ट्रेशन में गुणवत्ता मापदण्ड में योग्य नहीं पाया गया। point no. 5 door operated illumination lamp is fitted inside the chamber for easy visibility- not available and point no. 9 volt meter not available on panel.
		Deep Freezer-Vertical	Responsive	--
2.	M/s National Enterprises, Plot no. 138, DIC Industrial Area, Baddi, Solan, H.P. 173205	Deep Freezer – Horizontal & Deep Freezer-Vertical	Non-Responsive	1- दस्तावेजी मूल्यांकन किये जाने पर फर्म की निविदा आईटम के लिए दस्तावेजी रूप से योग्य नहीं पाया गया क्योंकि EMD की राशि बैंकर चैक या Demand Draft के रूप में जमा न कराकर F.D.R. जमा कराई गई है जो मान्य नहीं है। टेण्डर फीस का Demand Draft MD, RMSCL के नाम प्रस्तुत नहीं किया गया। यह MD, RISL के पक्ष में प्रस्तुत किया है जो निगम में जमा नहीं हो पाया। यह mandatory document है। 2. दस्तावेजी रूप से नॉन-रेसपोन्सिव होने के कारण तकनीकी परीक्षण के लिए आमंत्रित नहीं किया गया।
3.	M/s Electrocraft (India) Pvt., Plot No. 35, Sector-II, The Vasai Taluka Co-op Industrial Estate, Gaurajpada, Vasai (E) 401208 Paighar Maharashtra	Deep Freezer – Horizontal	Non-Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा कोटेड आईटम का डेमोस्ट्रेशन के लिए तकनीकी समिति के समक्ष उपस्थिति नहीं हुए।
		Deep Freezer-Vertical	Responsive	--
4.	M/s Authentic Instruments, 31, RIICI Indl. Area, Jhotwara, kamani Choraha, Jaipur	Deep Freezer – Horizontal	Responsive	--
		Deep Freezer-Vertical	Responsive	--

5.	M/s Remi Electrotechnik Limited. Remi House, 11 Cama Indl. Estate, Goregaon (E), Mumbai	Deep Freezer – Horizontal	Non-Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा कोटेड आईटम डेमोस्ट्रेशन में गुणवत्ता मापदण्ड में योग्य नहीं पाया गया। point no. 5 door operated illumination lamp is fitted inside the chamber for easy visibility- not available and point no. 9 volt meter not available on panel.
		Deep Freezer-Vertical	Non-Responsive	तकनीकी परीक्षण के दौरान फर्म द्वारा कोटेड आईटम डेमोस्ट्रेशन में गुणवत्ता मापदण्ड में योग्य नहीं पाया गया। point no. 5 door operated illumination lamp is fitted inside the chamber for easy visibility- not available.

Online financial bids of responsive bidders for above item will be opened on dated 07.03.2018 at 11.00 A.M.  
Demonstration report can be obtained from office during office time.



Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd.  
Jaipur

