


SPECIFICATIONS OF LARYNGOSCOPE SET, NEONATE

1. Laryngoscope set with Pediatric handle & blades.
2. Constituted of large hollow, cylindrical, slightly ribbed handle and set of depressors in stainless steel.
3. Handle is made of either chromium plated or stainless steel and can be opened at an extremity to insert two leak proof alkaline batteries (size AA Pencil Cell). The other end should have a stud contact which fits the various sizes and type of depressors.
4. The laryngoscope should be supplied with 02 Curved Depressors (Macintosh Type- size 1) - approximately 68 mm length (inside) with halogen bulb.
5. The laryngoscope should be supplied with 02 Straight Depressors (Miller Type- sizes 0 & 1 each) - approximately 53 mm length (inside) with halogen bulb.
6. The Laryngoscope set should be supplied in suitable protective good quality plastic box.
7. The manufacturer should be ISO 13485 / ISO 9001 certified and the Laryngoscope should be CE Certified.
8. The Laryngoscope Set should be supplied with following accessories:
 - a) 08 Spare halogen bulbs (02 for each depressor).
 - b) 04 Leak proof Alkaline Batteries (size AA Pencil Cell).
 - c) User manual with trouble shooting guidance (in English & Hindi).
 - d) Technical manual with maintenance and first line technical intervention instructions.
 - e) List of Service engineers in Rajasthan.
9. Guarantee: The Complete set should have guarantee of one year.
10. Consumables: Halogen Bulb & Battery only.



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur