


**TECHNICAL SPECIFICATION**  
**ECG MONITOR**

1. Should have at least 5" Color TFT Display.
2. Should have ECG as standard parameter.
3. Should monitor Lead I, II, III, aVr, aVl, aVf & V with help of 5 lead ECG cable.
4. Should have built-in rechargeable battery.
5. Should have battery backup of at least 90 minutes.
6. Should have 24 hours graphical and tabular trends.
7. Should have arrhythmia detection.
8. Should have facility to store ECG waveforms.
9. Should be CE certified or USFDA approved.
10. Guarantee: Two years on equipment from the date of installation and 06 months on accessories.
11. CMC: CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years.
12. Equipment should be supplied with compatible 5 lead ECG cable (2 Nos.) & Operating / user Manual (1 No.)
13. Rates of Accessories and Consumable: Rates of accessories and consumable like Battery, 5 lead ECG cable should be provided by bidder in BOQ in e-proc. Portal.
14. Installation will be done by supplier free of cost.
15. The service engineer should be based in Rajasthan.
16. Demonstration of Equipment is must for technical approval.
17. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.
18. Firm should mention all the pre-installation requirements in technical bid.

Delivery Period - 60 Days

  
**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation Ltd,**  
**Jaipur**