

TECHNICAL SPECIFICATION OF IUCD KIT

S. No.	Name of Item	Qty. in a set
1	Cheetal Forceps	1 No.
2	Cusco speculum 110 mm (L) X 35 mm (B)	1 No.
3	Sponge Forceps	1 No.
4	Artery Forceps 8" inch (Curved)	1 No.
5	Tenaculam (Single tooth) Double Curve 285 mm	1 No.
6	Uterine sound sims 330 mm	1 No.
7	Thread Cutting Scissor	1 No.
8	Vaginal Speculum Sims 15 cm (L) {34mm (B) X 30 mm (B)}	1 No.
General Terms & Conditions:		
9	All surgical instruments should be CE certified (valid documentation should be submitted in technical Bid)	
10	All surgical instruments should be manufactured by ISO 9001:2008 & ISO 13485:2003 (valid documentation should be submitted in technical Bid)	
11	Manufacturer should have ISO 7153 certificate. (valid documentation should be submitted in technical bid)	
12	Stainless Steel grade AISI 410 and 420, test report should be submitted.	
13	Laser marking along with code should be on the instrument.	
14	Guarantee: One year against any manufacturing defect.	
15	The company should mention the make & model of the quoted items and submit the technical compliance sheet in technical Bid	
16	Final technical approval after sample demonstration.	


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur