


TECHNICAL SPECIFICATION
STRETCHER ON TROLLEY

MS trolley frame powder coated with 175-225 mm dia. castors conforming to 4035 with latest amendment and removable stretcher frame size overall 1985 mm X 565 mm, made of MS pipe dia. 31.25, thickness 1 mm and bed made of MS sheet thickness 1.0 mm super imposed SS sheet of 0.8 mm thickness.

ch


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur