

TECHNICAL SPECIFICATION

WHEEL CHAIR

1. The wheel chair shall be made up of 16 gauge Stainless steel 304 grade tube frames and 16 gauge stainless steel 304 sheet for seat and backrest.
2. The dimension of wheel chair: 670mm (w) x 1120mm (d) x 92mm (h)±5mm.
3. It should have a fixed arm set with soft polyurethane.
4. It should have reticulated and breathable cushion and seat.
5. It should have retractable aluminium foot rest.
6. It should have push handle with plastic cap.
7. It should have 6" swivel nylon castor front wheel, 24" bicycle type rear wheel with puncture proof flat free type tyre.
8. The back wheel fixing bolt shall be covered with cup type nut.
9. It should have wheel breaking system on both side.

Other terms and conditions:

1. Manufacturer should be ISO- 9001 certified.
2. Guarantee: should be one year for all items from the date of successfully installation at consignees address.
3. Final Technical approval after demonstration.
4. Installation will be done by supplier free of cost.

The Company should submit technical compliance sheet as per technical specifications mentioning the make and model of quoted item along with catalogue in technical bid.



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd,
Jaipur