


## **SPECIFICATION OF BED SIDE SCREEN (3 PANEL)**

1. Made of M.S. Tubular frame in 3 folds.
2. Should be mounted on 5 cm antistatic castors with curtains wire on hooks.
3. Size: 66" H x 96" L (tolerance  $\pm 5\%$ )
4. Thickness of pipe 16 Gauge of 0.75" diameter.
5. Pre-treated and epoxy powder coated.
6. The screen cloth should be green or blue colour and washable.

### **General Terms & Conditions:**

1. **Guaranty:** Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.



Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd.,  
Jaipur

