


SPECIFICATION OF PATIENT STRETCHER TROLLEYS

1. Trolley made of pipe frame M.S. Tube 16 Gauge 1" round for horizontal & 1.25" for vertical.
2. Standard length & width and height 210 (L) X 65 (W) X 65-95 (H) cm
3. Should have four noiseless 15 cm diameter heavy duty castor wheels, 2 Nos. with locks for break movement.
4. Wheels should not be welded out the trolley as repair / replacement of wheels later on is possible.
5. The trolley should have facility to keep Oxygen cylinder.
6. Should have weight bearing capacity of around at 200 Kg (Patient + O2 cylinder).
7. Head-end of the trolley (around 2.25 + 2.5 feet) should be adjustable to put at deferent angles i.e.15 to 75 degree.
8. Detachable top should have this adjustable head end.
9. Arrangement for I/V stand and for supporting O2 (oxygen) cylinder.
10. 1.5" mattress at top, density of foam should be ≥ 40 .
11. Pre-treated and epoxy powder coated finished.

General Terms & Conditions:

1. Guaranty: Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.



Executive Director (EPM)

**Rajasthan Medical Services Corporation Ltd.,
Jaipur**

