



Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmse-rj@nic.in

CIN : U24232RJ2011SGC035067

Website: <http://rmse.health.rajasthan.gov.in>

F-80) RMSC/EPM/M-5/NIB-857/2025-26/RC-879/596

Dated: 02/07/2025

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals
Director (PH/HA/IEC/AIDS/RCH)

Sub: - Approval/Purchase of item "BIPAP Ventilator with well cushioned face mask and harness"

Please find enclosed herewith approved firm and item 'BIPAP Ventilator with well cushioned face masks and harness' for the rate contract and supply ending on 30.06.2027. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., website www.rmse.health.rajasthan.gov.in

In reference to the above **Rate contract No. 879** the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur..

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

F-80) RMSC/EPM/M-5/NIB-857/2025-26/RC-879/596

Dated: 02/07/2025


Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Jaipuria Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zanana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital- Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.02 10:16:00 IST
Reason: Approved






19. Deputy Director (IT), RMSC, Jaipur
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s ResMed India Pvt. Ltd., Cowrks Aerocity, First Floor, World mark1, Asset Area 11, Aerocity Hospitality District, IGI Airport, New Delhi-110037 Email: enquiries@resmed.co.in
22. Guard file.

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd, Jaipur

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.02 10:16:00 IST
Reason: Approved



Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: <http://rmisc.health.rajasthan.gov.in>

F-80) RMSC/EPM/M-5/NIB-857/2025-26/RC-879/ 596


Dated: 02/07/2025

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	BIPAP Ventilator with well cushioned face mask and harness
2.	Name & Address of Firm	M/s ResMed India Pvt. Ltd., Cowrks Aerocity, First Floor, World mark1, Asset Area 11, Aerocity Hospitality District, IGI Airport, New Delhi-110037 Email: enquiries@resmed.co.in
3.	Bidder Status	Importer
4.	Performance Security	Performance Security Deposit through BG No. PEBNDH297297 Dated 03-06-2025 HSBC Bank Valid up to 31.12.2028 Rs. 10,08,000/-
5.	Brand/Model	Resmed/ Lumis 150 VPAP St-A
	HSN Code	90192090
7.	Packing Unit	Each
8.	R/C No. & Validity	R.C. No. 879 (Valid Up to 30-06-2027)
9.	Indicative Quantity	200
10.	Rates for supply to RMSC from Suppliers	
	(i) Basic Rate (in Rs.)	90000.00
	(ii) IGST @ 12% (in Rs.)	10800.00
	(iii) CGST (in Rs.)	0.00
	(iv) SGST(in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	100800.00
16.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	90000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	4500.00
	(iii) CGST@ 6% { On (i)+(ii)} (in Rs.)	5670.00
	(iv) SGST@ 6% { On (i)+(ii)} (in Rs.)	5670.00
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	105840.00
22.	TECHNICAL SPECIFICATIONS of BIPAP Ventilator with well cushioned face mask and harness	
	Sr. No.	Technical Specification
	1	Non-invasive ventilation support system to ensure proper ventilation to COPD patients.
	2	It should be suitable for Adult and Pediatric patients.
	3	Modes of operation: CPAP, Spontaneous, S/T, Time Mode & Pressure Control.
	4	It should have display to monitoring facility for Tidal Volume, RR, Minute ventilation, I: E ratio.
	5	Operating pressure range of IPAP : 4 to 30 cm H2O or more. EPAP : 4 to 20 cm H2O or more.

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.02 10:16:00 IST
Reason: Approved



6	CPAP : 4 to 20 cm H2O or more. Unit should be capable of delivering Tidal Volume 1200 ml or more and have option to deliver guaranteed tidal volume.
7	Breath rate should be able to set up to 30 BPM or more with Ti of 0.5 to 3 sec. and rise time 100ms to 600ms.
8	Sensitivity settings: should have automatic trigger, cycle setting and leak compensation.
9	Unit should have integrated alarm for power, Pressure (High/Low), Low Minute ventilation, Apnea, Circuit disconnected etc. With digital technology for improved leak compensation during inspiration & exhalation.
10	Complete unit should be supplied in good quality bag which can carry adult and pediatric circuits with Adult and Pediatric masks along with equipment.
11	Quoted Model should have USFDA approval / CE certificate issued by a notified body registered in the European commission, under relevant category/BIS. Related documents / certification shall be submitted along with technical bid.
12	Quoted Model should have Valid CDSCO Certified related documents shall be submitted along with technical bid.
13	Unit should have battery backup capacity of 2 hour or more.
14	Unit should be supplied with reusable adult and pediatric patient circuits and reusable full face mask with harness Adult & Pediatric (1 no. each), operational / training manual (1 no.).
15	Guarantee: Three years from the date of installation.
16	CMC: CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
17	Installation will be done by supplier free of cost.
18	Firm should have service engineer network in Rajasthan.
19	The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.
23.	Cushioned face mask with harness Approved Rate: Basic rate 1100/- + GST@12%= Rs. 1232/- Each

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSC, Jaipur

Dated: 02/07/2025


F-80 RMSC/EPM/M-5/NIB-857/2025-26/RC-879/596

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zanana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.02 19:16:00 IST
Reason: Approved



12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s ResMed India Pvt. Ltd., Cowrks Aerocity, First Floor, World mark I, Asset Area 11, Aerocity Hospitality District, IGI Airport, New Delhi-110037 Email:enquiries@resmed.co.in
22. Guard file.

**Executive Director (EPM)
RMSC, Jaipur**

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.02.16:16:00 IST
Reason: Approved

