



Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065
CIN : U24232RJ2011SGC035067

E-mail: edepmrmse-rj@nic.in
Website: <http://rmsc.health.rajasthan.gov.in>

F-80 RMSC/EPM/M-5/NIB-845/2025-26/RC-883/705

Dated: 25/07/25

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals
Director (PH/HA/IEC/AIDS/RCH)

Sub: - Approval/Purchase of item "Multi Para Monitor (Basic)"

Please find enclosed herewith approved firm and item 'Multi Para Monitor (Basic)' for the rate contract and supply ending on 31.07.2027. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., website www.rmsc.health.rajasthan.gov.in

In reference to the above Rate contract No. 883 the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur..

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

F-80 RMSC/EPM/M-5/NIB-845/2025-26/RC-883/705

Dated: 25/07/25

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Jaipuria Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zanana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admin.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. Deputy Director (IT), RMSC, Jaipur

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.25 14:34:58 IST
Reason: Approved





20. All Zonal BMI are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s Allied Medical Limited, 76, Udyog Vihar, Phase IV, Gurugram-122015, Haryana, E-Mail-sales@alliedmed.co.in
22. Guard file.

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd, Jaipur

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.25 12:34:58 IST
Reason: Approved



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN: U24232RJ2011SGC035067

Website: <http://rmisc.health.rajasthan.gov.in>

F-80) RMSC/EPM/M-5/NIB-845/2025-26/RC-883/705

Dated: 25/07/2025

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Multi Para Monitor (Basic)
2.	Name & Address of Firm	M/s Allied Medical Limited 76, Udyog Vihar, Phase IV Gurugram-122015, Haryana E-Mail-sales@alliedmed.co.in
3.	Bidder Status	Manufacturer (MSME)
4.	Performance Security	Performance Security Deposit through BG No. 02611GP251101871 Dated 08-07-2025 for Rs.15,44,610/- Valid up to 31.12.2028, Kotak Mahindra Bank Ltd.
5.	Brand/Model	Allied/ Meditech M751
	HSN Code	90189019
7.	Packing Unit	Each
8.	R/C No. & Validity	R.C. No. 883 (Valid Up to 31-07-2027)
9.	Indicative Quantity	1500
10.	Rates for supply to RMSC from Suppliers	
	(i) Basic Rate (in Rs.)	91941.00
	(ii) IGST @ (in Rs.)	0.00
	(iii) CGST @ 6% (in Rs.)	5516.46
	(iv) SGST @ 6% (in Rs.)	5516.46
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.) Round of	102974.00
16.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	91941.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	4597.05
	(iii) CGST@ 6% { On (i)+(ii)} (in Rs.)	5792.28
	(iv) SGST@ 6% { On (i)+(ii)} (in Rs.)	5792.28
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	108122.61
22.	TECHNICAL SPECIFICATIONS OF Multi Para Monitor (Basic)	
	<ol style="list-style-type: none"> 1) It should having ECG, SpO2, NIBP, Respiration and two Temperature as standard parameters. 2) Should have facility to display 5/7 lead ECG, RR, HR, SpO2, NIBP, and dual Temperature. 3) Should display 6 or more waveforms of selected parameters simultaneously, should have colour coding for different waveforms. 4) Should have inbuilt battery backup of at least three hours. 5) Display: Color LED/LCD touch screen display size of 12" or more. 6) Monitor should have both touch screen and knob to operate. 7) Should have dual temperature monitoring in Celsius and Fahrenheit. 	

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.25 12:34:58 IST
Reason: Approved



- 8) Should have facility for displaying multi - screen configurations.
- 9) Should be able to store & display at least 4- days of graphical trends of all parameters.
- 10) Should be suitable for monitoring adult & paediatric & neonate patients.
- 11) The SpO2 technology should be Masimo/Nellcor and it should also sense in hypotension, shivering & motion. Tender Specific Authorization by OEM for Masimo/Nellcor technology and original sensor, should be attached with the technical bid.
- 12) Bill of entries/Chain documents of the purchase from Masimo/Nellcor must be submitted with invoice also at the time of bill submission.
- 13) Should have oscillometric technology for measurement of NIBP with Auto, START and Manual modes.
- 14) The pulse rate should be displayed either with ECG or SpO2.
- 15) The respiration rate should be calculated through Impedance method.
- 16) Should be able to analyze 20 or more type arrhythmias & ST segment changes and QT/QTc analysis.
- 17) Should have heart rate variability (HRV) analysis.
- 18) Should have defibrillator synchronization/Protection facility.
- 19) Memory should not wipe off when the power is turned off.
- 20) Should be able to give visual & audible alarms with three levels of volume adjustment. Monitor should have audio-visual alarm with snapshot (automatic & manual) and parameter lamp glowing in different colours.
- 21) Should have facility of connecting to the central monitoring station (Port only) having facility for real time monitoring through Ethernet cable or Wireless Connectivity.
- 22) Monitor should have VGA port/USB/HDMI or equivalent for External Display.
- 23) Wall mount stand should be provided free of the cost with the monitor.
- 24) Quoted model should be European CE (Notified Body)/USFDA approved/BIS. In case of Model not mentioned in certificate authenticity proof of quoted model from certifying agency shall be submitted with bid document.
- 25) Quoted model should be valid CDSO approved.
- 26) Warrantee-Three years on equipment from the date of installation.
- 27) CMC: CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Warrantee period of three years.
- 28) Equipment shall be supplied with compatible 5 lead ECG cable (1 Nos.), Masimo/Nellcor SpO2 (Adult, Pediatric, Neonate) probe with extension cable (Each 1No.), NIBP Cuff with extension cable of Large Adult, Adult & Paediatric (2 each), Temperature probe (2 Nos.) & Operating / user Manual (1 No.).

• The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSCL, Jaipur

Dated: 25/07/2025

F-80 RMSCL/EPM/M-5/NIB-845/2025-26/RC-883/705

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NIIM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(EST).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.

Signature valid

Digitally signed by Sandeep Charan
Designation: Executive Director
Date: 2025.07.25 12:34:58 IST
Reason: Approved



9. Principal, RUIIS College of Medical Sciences, Jaipur.
10. Superintendent Jaipuria Hospital Jaipur (Associated with RUIIS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zanana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RII)/(Hospital Admin.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s Allied Medical Limited, 76, Udyog Vihar, Phase IV, Gurugram-122015, Haryana, E-Mail- sales@alliedmed.co.in
22. Guard file.

Executive Director (EPM)
RMSC, Jaipur

Signature valid

Digitally signed by Surdeep Charan
Designation: Executive Director
Date: 2025.07.25 12:34:58 IST
Reason: Approved