

Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
 CIN:U24232RJ2011SGC035067

E-mail : edepmmrsc-rj@nic.in
 Website: http://rpsc.health.rajasthan.gov.in

No. F-8(RMSC/EPM/M-6//NIB-530/2022-23/320

Dated 20/7/2022

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	High Frequency X-Ray Machine (100mA)
2.	Name & Address of Firm	Epsilon Healthcare Solutions (P) Ltd G1-57 ,RIICO Indus.Area Phase-II, Bagru,Jaipur-303007 Rajasthan (India) E-mail: marketing@epsilonhealthcare.com
3.	Bidder Status	Manufacturer
4.	Performance Security	BG No:- 0012NDDG00025323 D.t. 04.07.2022 Amount- 3,20,320 (ICICI Bank)
5.	Brand/Model	Epsilon EP Corsa 6R
	HSN Code	90221490
6.	Packing Unit	Each
7.	R/C No. & Validity	R.C. No 768 (Valid Up to 30 June 2024)
8.	Indicative Quantity	130
9.	Rates for supply to RMSC from Suppliers	
	(i) Basic Rate (in Rs.)	440000.00
	(ii) IGST @ 12% (in Rs.)	52800.00
	(iii) CGST (in Rs.)	0.00
	(iv) SGST(in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	492800.00
10.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	440000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	22000.00
	(iii) CGST@6% { On (i)+(ii)} (in Rs.)	27720.00
	(iv) SGST@ 6% { On (i)+(ii)} (in Rs.)	27720.00
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	517440.00
Technical Specifications of Radiology High frequency X ray Machine (100 mA)		
01.	High frequency X-Ray machine suitable for general radiography.	
02.	X-RAY GENERATOR: • High Frequency generator of 50 KHz or more.	
03	• Radiographic KV: 40 to 100KV or more.	

[Handwritten Signature]



04.	• Output power: 6.0 KW	
05.	• mAs range: 100 mA or more.	
06.	• Exposure Time (Rad) 7ms to 7 sec.	
07.	CONTROL: • Two point selection technique i.e. KV and mAs.	
08.	• KV Increase & Decrease Switches.	
09.	• mAs Increase & Decrease Switches	
10.	• Machine ON/OFF Switch.	
11.	• Collimator Lamp 'ON' Switch -on collimator side and control board.	
12.	• Standby & Exposure Switch.	
13.	• Self diagnostic Programme with indicators for:-	
14.	• Earth fault Error	
15.	• KV Error/ Inverter error	
16.	• Filament Error	
17.	• Tube head Thermal Error	
18.	• Stand by (Ready) & X-Ray On Indicator.	
19.	• Incoming Voltage Indicator. There should be provision for the machine to work from 195Volts Input supply to 265V input supply.	
20.	• Anatomical Programming Radiography (i.e. APR) should be provided in which KV and mAs are automatically selected depending upon the physique of the patient and part of the body to be X-Rayed.	
21.	• Anatomical Programming of minimum 200 programmes should be provided.	
22.	• There should be a provision that the control should get off, if no key is pressed for 10 Min.	
23.	• There should be inbuilt delay between the exposures to protect x-ray tube from over loading.	
24.	• A Hand Switch with Dual action for exposure Release with Retractable Cord is provided for Radiation Protection to the Operator.	
25.	X-RAY TUBE: • One no. Dual focus Rotating Anode X-Ray tube thermally protected.	
26.	• Anode heat storage capacity of tube should be 100 KHU or more.	
27.	• One Pair of 8 meter H.V. Cable.	
28.	• One No. Collimator with auto shut off facility should be provided.	
29.	HV TANK: • A very compact H.V. Tank filled with high dielectric transformer oil or Monoblock HT tank should be provided. The H.V. Tank should contain H.V. transformer, Filament Transformers, H.V.	

af

[Signature]



55.	of CMC may be awarded for seven years after guarantee. 9. Installation will be done by supplier free of cost.	
56.	10. The Company should submit original catalogue and data sheet (signed & stamped) of quoted item at the time of bid submission. Any declaration/ undertaking in this regard shall not be considered.	
57.	11. The Firm shall provide guideline and support to consignee, required to register online official Website eLORA(e-Licensing of Radiation Application).	
58.	12. For item Radiology High Frequency X ray Machine (100 mA), it is mandatory for conducting a physical/virtual demonstration of quoted equipment at any site at their own (Firm) cost.	
59.	13. The supplier/ Manufacturer should give undertaking regarding the availability of spare parts for 10 years or more from the date of installation.	

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Bani

20/7/22
Executive Director (EPM)
RMSC, Jaipur

No. F-8()RMSC/EPM/M-6/NIB-530/2022-23/320

Dated 20/7/2022

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/ PHS/ Secy., Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to ACS/ PHS/ Secy., Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All CM&HO's., All P.M.O.'s. and All Dy. CM&HO's, Rajasthan.
17. Joint Director (IT), for uploading in e-tmts software and RMSCL website both.
18. All BME, RMSC
19. M/s Epsilon Healthcare Solutions (P) Ltd
20. Guard file.

20/7/22
Executive Director (EPM)
RMSC, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmse-rj@nic.in

CIN : U24232RJ2011SGC035067

Website: <http://rmse.health.rajasthan.gov.in>

No. F-8()RMSC/EPM/M-6//NIB-530/2022-23/321

Dated 20/7/2022

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, High Frequency X-Ray Machine (100mA) , for the Rate Contract and supply, ending on 30.06.2024.

Please find enclosed herewith approved firm and High Frequency X-Ray Machine (100mA), for the rate contract and supply ending on 30.06.2024. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmse.health.rajasthan.gov.in

In reference to the above rate contract No. F-8()RMSC/EPM/M-6//NIB-530/2022-23/..... dated..... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

Dated 20/7/2022

No. F-8()RMSC/EPM/M-6//NIB-530/2022-23/ 321

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/ PHS/ Secy., Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to ACS/ PHS/ Secy., Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NIIM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All CM&HO's., All P.M.O.'s. and All Dy. CM&HO's, Rajasthan.
17. Joint Director (IT), for uploading in e-tmts software and RMSCL website both.
18. All BME, RMSC
19. M/s Epsilon Healthcare Solutions (P) Ltd
20. Guard file.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur