

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmisc-rj@nic.in

CIN : U24232RJ2011SGC035067

Website: http://rmisc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-1/NIB-476/2021/ 6542

Dated: 10/03/2021

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Blood Bank Refrigerator (300 Blood Bag Capacity)', for the Rate Contract and supply, ending on 28.02.2023.

Please find enclosed herewith approved firm and Blood Bank Refrigerator (300 Blood Bag Capacity), for the rate contract and supply ending on 28.02.2023. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmisc.health.rajasthan.gov.in

In reference to the above rate contract No. F-8() RMSC/EPM/M-1/2020-21/NIB-476/6543...Dated..10/3/2021.. the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

No. F-8() RMSC/EPM/M-1/NIB-476/2021/ 6542


Dated: 10/03/2021

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Remi Elektrotechnik Limited, Mumbai.
22. Guard file.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

**Rajasthan Medical Services Corporation Limited, Jaipur****Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005**

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmsc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-1/NIB-476/2021/ 6543

Dated: 10/03/2021

Rate Contract and Price Details


S.N.	Item	Detail
1.	Name of Equipment / Item	Blood Bank Refrigerator (300 Blood Bag Capacity)
2.	Name & Address of Firm	M/s Remi Elektrotechnik Limited, Remi House, 11 Cama Indl. Estate, Goregaon (E), Mumbai-400 063, India E-Mail:- bids@remilabworld.com Contact No. : 9414522366
3.	Bidder Status	Manufacturer
4.	Performance Security	SBI BG No. 6122421BG0000296 dated 02.03.2021 Rs. 327500.00 (Expiry dt 31.03.2024)
5.	Brand/ Make	Remi
	Model	BR300 Ultra with voltage stabilizer
6.	Packing Unit	Each
7.	R/C No. & Validity	R.C. No. 664 (Valid Up to 28.02.2023)
8.	Indicative Quantity	60
9.	Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)	185000.00
	(ii) IGST @ 18% (in Rs.)	33300.00
	(iii) CGST (in Rs.)	0.00
	(iv) SGST (in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	218300.00
10.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	185000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	9250.00
	(iii) CGST@ 9% { On (i)+(ii)} (in Rs.)	17482.50
	(iv) SGST@ 9% { On (i)+(ii)} (in Rs.)	17482.50
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	229215.00
11.	Technical Specification of Blood Bank Refrigerator (300 Blood Bag Capacity) -	
	1. Purpose of Equipment	
	1. A refrigerator for storing whole blood.	
	2. Must be designed specifically for blood bank use. Commercial or modified commercial refrigerators for other purpose are not acceptable.	
	2. Type of Equipment	
	Approved standard electrical Blood bank refrigerator that uses a compressor circulating CFC-free refrigerant / green gas.	
	3. Quality Standard	
	1. Should have built in temperature recorder and controller positioned at Eye level for better visibility and	



monitoring.

2. Must have uninterrupted digital electronic temperature circular chart recorder with 21 days days memory with print facility / six inch, 24×7-day, ink-less pressure sensitive circular chart recorder and accuracy should be +/- 1 degree Celsius with supply of free charts for a period of Guaranty.
 3. Should have a digital temperature probe/RTD located in the top portion of the chamber in a liquid medium bottle for accurate temperature measurement.
 4. Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid).
 5. Quoted model should be compliant with European CE (Notified Body)/US FDA/BIS (Valid documentation should be submitted in technical bid).
 6. Equipment must meet electrical safety specifications of IEC 60601-1.
4. Capacity
1. At least 300 Whole Blood standard blood bags/250 double bag with Whole Blood standard blood bags.
5. Construction
1. Outside C. R. (Corrosion Resistant) Sheet at least 1 mm thick.
 2. Inside stainless steel of at least 22 G.
 3. Insulation 50-70 mm thick, foaming agent CFC free.
 4. Glass door with full visibility of unit Door without door opening. Thickness of door should be 50-70 mm.
6. Drawers
1. Stainless steel, scratch resistant Drawers perforated to ensure good air circulation.
 2. Roll out type.
 3. At least four or more in number.
7. Door
1. Glass door with full visibility of units without opening door.
 2. Automatic/Magnetic closing at 90° closer.
 3. Door opening audio and visual alarm.
 4. Door lock should be available.
8. Electrical characteristics
1. Compatible with Input voltage: 240V 50 Hz Single phase AC.
 2. Equipment shall be supplied with external Servo Voltage Stabilizer with input voltage rating 90-280V and 50Hz.
 3. The Compressor running time should be maximum of 35%.
 4. At room temperature of 25°C should be able to maintain at ideal compressor running time of 27%.
9. Internal Temperature
1. Blood Bank Refrigerator should have inside temperature range of 2°C - 6°C.
 2. User parameter settings: Set point, High alarm point, low alarm point, buzzer off time, C/F unit display choice.
 3. Whatever the load, setting accuracy less than or equal to 0.5°C (preferably 0.1°C).
 4. Should ensure frost free performance thereby avoiding either freezing or heating. If defrosting function used, temperature should not go outside range specified above.
 5. Four hours battery backup for thermographic recording (Chart Recorder) display and alarm.
10. External Ambient Temperature:
1. Can perfectly maintain internal temperature as above at full load in an ambient temperature of +10 to at least +40 °C (± 2°C).
11. Hold-Over Time
1. A full load (At least 300 Whole Blood standard blood bags/250 double bag with Whole Blood standard blood bags) at +4 °C (±1 °C) should take * more than 1.5 hours to rise to above +6 °C if power off without opening the door at room temperature 22°C(±2 °C).

10/3/20



12. Cooling Down Time.

1. A full load of blood packs at +25 °C takes a maximum of 6 Hrs for all packs to reach +6 °C.
13. Should have flicker free CFL lamp for uniform lighting and better visibility of samples inside the cabinet.
14. Should have Audible and visual High and Low temperature alarm & power fail alarm as a standard feature with alarm silence button.
15. Provision to connect with central (Temperature) monitoring system / facility of wireless data transfer facility & storage of temp. data on cloud, with continuous accessibility of data on PC or laptop and on smart phones with internet facility.
16. Guarantee: Three years on equipment from the date of installation.
17. CMC: CMC will be given @ 5 % of net rate - exclusive GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.
18. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
19. Installation will be done by supplier free of cost.
20. The service engineer should be based in Rajasthan.
21. Demonstration of equipment is must for final technical approval.

Other terms and condition to be covered during guarantee period by the supplier:-

- Response time- < 48 Hours after logged the complain in complain portal.
- Service hours- As per health facility schedule.
- Preventive Maintenance Schedule (PM)-As per OEM (Mention PM Schedule).
- Calibration Schedule – As per OEM (Mention Schedule).
- Up time- 95% (346 Days).
- Technical & Application Support, Demonstrations & Trainings - As & when required.
- Toll Free No-
- Life of the equipment- As per OEM.

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSC, Jaipur
Dated: 10/03/2021

No. F-8() RMSC/EPM/M-1/NIB-476/2021/ 6548

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)



15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. **M/s Remi Elektrotechnik Limited, Mumbai.**
22. Guard file.

Executive Director (EPM)
RMSC, Jaipur