



Rajasthan Medical Services Corporation Limited

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur – 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmssc.health.rajasthan.gov.in

F-8(218) RMSC/EPM/M-2/2020-21/NIB-444/ 182

dated: 03/11/2020

Order for Rate Contract

Item/Bidder status: Importer

Indicative Quantity: 04

S. No.	Item	Detail
1.	Name of Equipment / Item	FTIR-Spectrophotometer
2.	Name & Address of Firm	M/s Ginger Science Pvt. Ltd. LGF6-8, Jinkushal Suri Complex, 230-A, Masjid Moth, South Ex-II, New Delhi-110049 Email: sales@gingerscience.co.in
3.	Make	ThermoFisher Scientific
4.	Model	Nicolet Summit
5.	Packing Unit	Each
6.	R/C No. & Validity	(R.C. No 615 Valid upto 30.09.2022)
7.	Net Rate Per packing Unit (Rs.)	19,50,000.00 (Rs. Ninteen Lac Fifty Thousand Only) HSN Code 90273090
8.	RMSC Surcharge @5%	Rs. 97500.00
9.	Rates for Supply to RMSC from Suppliers	
(i)	Basic Rate (In Rs.)	1950000.00
(ii)	IGST@ 18 % (In Rs.)	351000.00
(iii)	CGST @ 0 % (In Rs.)	0.00
(iv)	SGST @ 0 % (In Rs.)	0.00
(v)	Approved Rate for Suppliers From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	2301000.00
10.	Rate for Supply To Consignee From RMSC	
(i)	Basic Rate (In Rs.)	1950000.00
(ii)	RMSC Surcharge @ 5% on (In Rs.)	97500.00
(iii)	CGST @ 9% {On (i)+(ii)} (In Rs.)	184275.00
(iv)	SGST @ 9% {On (i)+(ii)} (In Rs.)	184275.00
(v)	Approved Rate for Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	2416050.00
11.	Technical specifications of - FTIR-Spectrophotometer is :-	
	Fully Computer Controlled COMPACT BENCH-TOP FTIR system with universal sample compartment.	
1	Sample module must be automatically identified.	
2	The system should have latest digital signal processor.	
3	The system should indicate whether the source and laser are operational.	
4	The System should have feature for humidity and vapour protection.	
5	Wave number range: 6000 to 500 cm-1	
6	Source: Long Life IR Source	
7	Detector: MID-IR /DLTGS/DLATGS/LiTaO3 detector with temperature control mechanism	
8	Resolution: 0.6 cm ⁻¹ or less	
9	Signal to Noise Ratio- 25000:1 Or better peak to peak for 1 min.	
10	Wave number precision: 0.05 cm-1 or better at 2000 cm-1	
11	Beam splitter: KBr coated with Ge/ZnSe/CaF2	
12	The software should also have: -Compare Software, Spectral Search; Quantitative Analysis software Spectral interpretation for unknowns; Quality checks programs and CFR-21 Part-11 Compliance should be available.	
13	The software should have real time data collection and should have the facility to continuously monitor the performance of source, detector, power supply and laser.	

Handwritten signature and date: 03/11/2020



14	Libraries: Built in Library with Minimum 1000 reference spectra for pharmaceutical products and drugs.
15	ZnSe- Attenuated Total Reflection (ATR) -1 Nos.
16	IR Grade KBr – 3x100 gm
17	Fixed Volume Liquid Cell and fixed thickness (0.5 mm) -1 Nos.
18	Sodium Chloride pellets with Holder for liquid paraffin mulls-1 No.
19	Hydraulic Press of suitable capacity
20	Set of 13 mm KBr die, Pellet Holder and Nozzle oil should be supplied-2 Nos.
21	Polystyrene film of 0.3 mm having NIST Traceable certificates with atleast five year validity - One
22	Agate Mortar Pestle (Dia 2 inch) -2 Nos.
23	<u>Computer Specifications: (Higher configuration are acceptable)</u> <u>Make :HP/Dell/IBM</u> Processor - Intel i5 (5 th generation) ; RAM - 4 GB ; Hard disk - 1 TB ; Graphic Card; DVD writer; 19 ” TFT screen ; LAN Port ; USB Ports ; Wi-Fi ; Multimedia Keyboard ; Optical Mouse Operating System – Preloaded Windows 10 Pro ; Antivirus Printer- Laser Printer Monochrome with duplex printing and LAN port. UPS-2 KVA, Single phase with 60 minutes backup.
24	Documents /Training and other notes: (i) The equipment should be certified by European CE marking or UL marking. (ii)Should have safety certificate from a competent authority like CE/Equivalent or valid detail of electrical and functional safety test. Copy of the certificate/test report shall be produced along with the technical bid (iii)Comprehensive guarantee period with spare parts for at least three years from the date of installation. (iv)Instruction Manuals (Hard Copy or Soft copy as applicable) along with standard Operation Procedures (SOP) and appropriate Dust cover for the instrument. (v)Installation qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) documents should be provided at the time of supply of the instrument. (vi)Satisfactory technical and application training to the personals at the site after installation. (vii)CMC will be given @5% of net rate-inclusive of GST (as applicable) & yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion guarantee period. (viii)Minimum of two routine maintenance services/inspections to be provided annually and any number of breakdown emergency calls should be attended during the guarantee period. (ix)Price quoted should be inclusive of their complete installation in all respects at site as per purchase order. (x)Final technical approval only after demonstration of the quoted product. (xi)Any other perquisite required (if any) with specifications should be included with the instrument to run and complete the installation.


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur

F-8(218) RMSC/EPM/M-2/2020-21/NIB-444/ 182

dated: 03/11/2020

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to PHS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)



15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur.
20. M/s Ginger Science Pvt. Ltd. LGF6-8, Jinkushal Suri Complex, 230-A, Masjid Moth, South Ex-II, New Delhi-110049 Email: sales@gingerscience.co.in
21. Guard file.

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur

