

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmisc-rj@nic.in

CIN : U24232RJ2011SGC035067

Website: <http://rmisc.health.rajasthan.gov.in>

No. F-80 RMSC/EPM/M-1/2020-21/NIB-469/ 7235

Dated : 27/05/2021

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Laminar Air Flow (Vertical)', for the Rate Contract and supply, ending on 31.05.2023.

Please find enclosed herewith approved firm and Laminar Air Flow (Vertical), for the rate contract and supply ending on 31.05.2023. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmisc.health.rajasthan.gov.in

In reference to the above rate contract No. No. F-80 RMSC/EPM/M-1/2020-21/NIB-469/..... dated..... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

No. F-80 RMSC/EPM/M-1/2020-21/NIB-469/ 7235

Dated : 27/05/2021

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s:
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s Authentic Instruments, 31, RIICO Industrial Area, Jhotwara, Near Kamani Chouraha, Jaipur-302012 Email: support@authenticinstruments.com, calibration@authenticinstruments.com
22. Guard file.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmsc.health.rajasthan.gov.in

No. F-80) RMSC/EPM/M-1/2020-21/NIB-469/ 7236

Dated : 27/05/2021

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Laminar Air Flow (Vertical).
2.	Name & Address of Firm	M/s Authentic Instruments, 31, RIICO Industrial Area, Jhotwara, Near Kamani Chouraha, Jaipur-302012 Email: support@authenticinstruments.com , calibration@authenticinstruments.com
3.	Bidder Status	Manufacturer (MSME)
4.	Performance Security	11,000 Through D.D. No. 030233/27.04.2021 HDFC Bank
5.	Make	Authentic
	Brand/Model	Laminar Air Flow(2x2x2)
	HSN Code	90189099
6.	Packing Unit	Each
7.	R/C No. & Validity	R.C. No 674 (Valid Up to 31 May 2023)
8.	Indicative Quantity	30
9.	Rates for supply to RMSC from Suppliers	
	(i) Basic Rate (in Rs.)	60000.00
	(ii) IGST @ 18% (in Rs.)	10800.00
	(iii) CGST (in Rs.)	0.00
	(iv) SGST(in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	70800.00
10.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	60000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	3000.00
	(iii) CGST@ 9% { On (i)+(ii)} (in Rs.)	5670.00
	(iv) SGST@ 9% { On (i)+(ii)} (in Rs.)	5670.00
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	74340.00
11.	Technical Specifications of Laminar Air Flow (Vertical)	
	<ol style="list-style-type: none"> Construction-fabricated from welded 16 gauge C.R. steel includes adjustable levelers Working area - 2 x 2 x 2 and it should be made from SS 304 grade. Finish: exterior surface feature a chip-resistant, powder-coated finish of approximately 0.6 mm thickness Front panel: movable transparent scratch resistant sheet of approx. 6mm thickness Side panels should be made of thick transparent Acrylic sheet. Pre-filters: disposable fiber glass with Filtration efficiency of 98% for all types of particles of size 8 micron and larger. 	

27/5/21



7. HEPA filter: choice of HEPA (rated 99.97% efficient @ 0.3µm particles) or ULPA (rated 99.99% efficient @0.12µm particles)
8. Lighting: Fluorescent type with lux > 800 & UV light.
9. On and OFF timer indicator for UV light along with total time usage of UV light
10. Blower Motor Assembly-Duly balanced, direct drive motor blower units sized to provide adequate air flow volume over the entire surface of HEPA Filter.
11. FFU airflow: 650 CFM (2" x 4"), 460 (2"x 3") of air at filter pressure drop from 0.35-0.795" W.G at 90 FPM.
12. Noise level should be less than 65 db.
13. Appropriate manometer to measure the pressure should be provided along the equipment.
14. Castor Wheels for easy movability.
15. Operates on 230 Volts AC Single phase 50 Hz.
16. On site annual calibration and validation.
17. On site counting of dust particles in laminar air flow chamber annually under guarantee
18. Installation, commissioning and trial run will be the responsibility of the supplier
19. The system should be CE approved (Valid documentation should be submitted in technical bid).
20. Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid)
21. Guarantee: Three years on equipment from the date of installation.
22. CMC will be given @5% of net rate-inclusive of GST (as applicable) & yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion guarantee period.
23. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
24. Installation will be done by supplier free of cost.
25. The service engineer should be based in Rajasthan.
26. Four Preventive Maintenance services annually during Guarantee and CMC period are essential.
27. Demonstration of equipment is must for final technical approval.

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSC, Jaipur

No. F-80 RMSC/EPM/M-1/2020-21/NIB-469/ 7236

Dated : 27/05/2021

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.



13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s Authentic Instruments, 31, RIICO Industrial Area, Jhotwara, Near Kamani Chouraha, Jaipur-302012 Email: support@authenticinstruments.com, calibration@authenticinstruments.com
22. Guard file.


Executive Director (EPM)
RMSC, Jaipur